



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

## SVOC DATA PACKAGE

### Client Project Information

Project ID:  
Project Description:  
Contact: Emily Jones

### ALSE Project Information

Project ID: FLS100  
Contact: Breanne Dusureault  
Submission ID(s): L2310263

Final Package Review by:

A handwritten signature in black ink, appearing to read "Breanne", is written over a horizontal line.

Date Reviewed: 23-Oct-19



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

## SVOC DATA PACKAGE

### SECTION 1: PROJECT NARRATIVE

#### ALSE Project Information

Project ID: FLS100

Contact: Breanne Dusureault  
Submission ID(s): L2310263

Analytical Method: PCDD/F by EPA TO9A

#### Client Project Information

Project ID:

Project Description:  
Contact: Emily Jones

ALS Sample ID	Client Sample Descriptions	Matrix	Date Sampled	Date Received	Date Extracted	Date Analyzed
L2310263-1	HEISER-25764410	Puf	11-Jul-19	16-Jul-19	12-Aug-19	16-Aug-19
L2310263-2	CITY-45764407	Puf	11-Jul-19	16-Jul-19	12-Aug-19	16-Aug-19
L2310263-3	RES-25764409	Puf	11-Jul-19	16-Jul-19	12-Aug-19	16-Aug-19
WG3116369-1	Method Blank	MEDIA	n/a	n/a	12-Aug-19	16-Aug-19
WG3116369-4	Method Blank	REAGENT	n/a	n/a	12-Aug-19	16-Aug-19
WG3116369-2	Laboratory Control Sample	QC	n/a	n/a	12-Aug-19	16-Aug-19

#### Comments and Notes:

##### a) Sample Integrity:

The samples were received in good condition at 6.8 degrees C.

##### b) Instrumental Analysis:

The recovery of the labelled standard 13C12-1,2,3,4,6,7,8-HpCDF was above the method control limit for the pre-run continuing calibration verification (CCV). The recoveries are within limits for the post-run CCV and the samples. However, the reported recoveries for this standard may be elevated. No bias to native target data is expected.

The recoveries of select field standards are below the method control limit.

I certify that this data package is in compliance with the terms and condition of the contract , both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this data package (hardcopy and/or electronic version) has been authorized by the Laboratory Manager or his designee, as verified by the following signature.

Steve Kennedy, Technical Supervisor

23-Oct-19

Date



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

**SVOC DATA PACKAGE**

**SECTION 2a:**

**CERTIFICATE OF ANALYSIS**

**SAMPLE AND QUALITY CONTROL SUMMARIES**



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6  
Phone: 905-331-3111, FAX: 905-331-4567

## Certificate of Analysis

<b>ALS Project Contact:</b>	Breanne Dusureault	<b>Client Name:</b>	Floyd Snider
<b>ALS Project ID:</b>	FLS100	<b>Client Address:</b>	601 Union Street, Suite 600
<b>ALS WO#:</b>	L2310263		Seattle, WA 98101
<b>Date of Report</b>	25-Oct-19		USA
<b>Date of Sample Receipt</b>	16-Jul-19	<b>Client Contact:</b>	Emily Jones
		<b>Client Project ID:</b>	

**COMMENTS:** **PCDD/F by EPA T09A**  
Analysis by HRGC/HRMS (modified EPA 1613B)

The recoveries of select field standards are below the method control limit.

Certified by:

A handwritten signature in black ink, appearing to read "Steve Kennedy", is written over a horizontal line.

Steve Kennedy  
Technical Supervisor

Results in this certificate relate only to the samples as submitted to the laboratory.

This report shall not be reproduced, except in full, without the written permission of ALS Canada Ltd.

# ALS Life Sciences

## Sample Analysis Summary Report

Sample Name	HEISER-25764410	CITY-45764407	RES-25764409
ALS Sample ID	L2310263-1	L2310263-2	L2310263-3
Sample Size	1	1	1
Sample size units	Puf	Puf	Puf
Percent Moisture	n/a	n/a	n/a
Sample Matrix	PUF	PUF	PUF
Sampling Date	11-Jul-19	11-Jul-19	11-Jul-19
Extraction Date	12-Aug-19	12-Aug-19	12-Aug-19
<b>Target Analytes</b>	<b>pg</b>	<b>pg</b>	<b>pg</b>
2,3,7,8-TCDD	<1.4	<2.1	<2.0
1,2,3,7,8-PeCDD	<0.61	<1.0	<0.89
1,2,3,4,7,8-HxCDD	<0.85	<1.2	<1.1
1,2,3,6,7,8-HxCDD	<0.78	<1.1	<1.0
1,2,3,7,8,9-HxCDD	<0.81	<1.1	<1.1
1,2,3,4,6,7,8-HpCDD	<1.3	<1.4	<2.4
OCDD	<6.5	8.24	<5.9
2,3,7,8-TCDF	<0.99	<1.2	<1.2
1,2,3,7,8-PeCDF	<0.65	<0.82	<0.93
2,3,4,7,8-PeCDF	<0.58	<0.73	<0.84
1,2,3,4,7,8-HxCDF	<0.65	<0.67	<0.70
1,2,3,6,7,8-HxCDF	<0.61	<0.62	<0.65
2,3,4,6,7,8-HxCDF	<0.65	<0.66	<0.69
1,2,3,7,8,9-HxCDF	<0.74	<1.1	<1.3
1,2,3,4,6,7,8,9-HpCDF	<0.55	<0.81	<0.89
1,2,3,4,7,8,9-HpCDF	<0.65	<0.84	<1.1
OCDF	<1.5	<1.3	<1.7
<b>Field Spike Standards</b>	<b>% Rec</b>	<b>% Rec</b>	<b>% Rec</b>
37C14-2,3,7,8-TCDD	82	80	87
13C12-1,2,3,4,7,8-HxCDD	78	78	80
13C12-2,3,4,7,8-PeCDF	79	80	91
13C12-1,2,3,4,7,8-HxCDF	58	56	70
13C12-1,2,3,4,7,8,9-HpCDF	69	61	75
<b>Extraction Standards</b>			
13C12-2,3,7,8-TCDD	70	92	64
13C12-1,2,3,7,8-PeCDD	82	119	81
13C12-1,2,3,6,7,8-HxCDD	84	108	82
13C12-1,2,3,4,6,7,8-HpCDD	81	120	76
13C12-OCDD	68	118	57
13C12-2,3,7,8-TCDF	86	102	85
13C12-1,2,3,7,8-PeCDF	93	120	89
13C12-1,2,3,6,7,8-HxCDF	88	110	86
13C12-1,2,3,4,6,7,8-HpCDF	90	118	83
<b>Cleanup Standard</b>			
13C12-1,2,3,7,8,9-HxCDF	86	117	89
<b>Homologue Group Totals</b>	<b>pg</b>	<b>pg</b>	<b>pg</b>
Total-TCDD	<1.4	<2.1	<2.0
Total-PeCDD	<0.61	<1.0	<0.89
Total-HxCDD	<0.85	<1.2	<1.1
Total-HpCDD	<0.75	<0.68	<2.4
Total-TCDF	<0.99	<1.2	<1.2
Total-PeCDF	<0.65	<0.82	<0.93
Total-HxCDF	<0.74	<0.76	<0.79
Total-HpCDF	<0.65	<0.84	<1.1
<b>Toxic Equivalency - (WHO 2005)</b>			
<b>Lower Bound PCDD/F TEQ (WHO 2005)</b>	0.00	0.00247	0.00
<b>Mid Point PCDD/F TEQ (WHO 2005)</b>	1.43	2.14	2.06
<b>Upper Bound PCDD/F TEQ (WHO 2005)</b>	2.84	4.14	3.99

# ALS Life Sciences

## Quality Control Summary Report

Sample Name	Method Blank	Method Blank	Laboratory Control Sample
ALS Sample ID	WG3116369-1	WG3116369-4	WG3116369-2
Sample Size	1	1	1
Sample size units	Puf	Puf	n/a
Percent Moisture	n/a	n/a	n/a
Sample Matrix	MEDIA	REAGENT	QC
Sampling Date	n/a	n/a	n/a
Extraction Date	12-Aug-19	12-Aug-19	12-Aug-19
<b>Target Analytes</b>	<b>pg</b>	<b>pg</b>	<b>% Rec</b>
2,3,7,8-TCDD	<2.3	<2.0	105
1,2,3,7,8-PeCDD	<1.3	<1.1	106
1,2,3,4,7,8-HxCDD	<1.4	<1.5	102
1,2,3,6,7,8-HxCDD	<1.3	<1.3	109
1,2,3,7,8,9-HxCDD	<1.3	<1.4	117
1,2,3,4,6,7,8-HpCDD	<2.3	<2.1	102
OCDD	<6.2	8.95	97
2,3,7,8-TCDF	<1.5	<1.4	102
1,2,3,7,8-PeCDF	<1.1	<0.82	109
2,3,4,7,8-PeCDF	<0.97	<0.74	101
1,2,3,4,7,8-HxCDF	<0.97	<0.79	102
1,2,3,6,7,8-HxCDF	<0.90	<0.73	115
2,3,4,6,7,8-HxCDF	<0.97	<0.78	111
1,2,3,7,8,9-HxCDF	1.11	<0.93	112
1,2,3,4,6,7,8-HpCDF	<1.2	<0.76	103
1,2,3,4,7,8,9-HpCDF	<1.4	<0.90	101
OCDF	3.42	2.42	98
<b>Field Spike Standards</b>	<b>% Rec</b>	<b>% Rec</b>	<b>% Rec</b>
37Cl4-2,3,7,8-TCDD	NS	NS	NS
13C12-1,2,3,4,7,8-HxCDD	NS	NS	NS
13C12-2,3,4,7,8-PeCDF	NS	NS	NS
13C12-1,2,3,4,7,8-HxCDF	NS	NS	NS
13C12-1,2,3,4,7,8,9-HpCDF	NS	NS	NS
<b>Extraction Standards</b>			
13C12-2,3,7,8-TCDD	48	52	69
13C12-1,2,3,7,8-PeCDD	60	68	93
13C12-1,2,3,6,7,8-HxCDD	56	59	77
13C12-1,2,3,4,6,7,8-HpCDD	59	67	100
13C12-OCDD	48	54	102
13C12-2,3,7,8-TCDF	67	67	80
13C12-1,2,3,7,8-PeCDF	68	73	97
13C12-1,2,3,6,7,8-HxCDF	69	62	82
13C12-1,2,3,4,6,7,8-HpCDF	71	74	104
<b>Cleanup Standard</b>			
13C12-1,2,3,7,8,9-HxCDF	66	70	89
<b>Homologue Group Totals</b>	<b>pg</b>	<b>pg</b>	
Total-TCDD	<2.3	<2.0	
Total-PeCDD	<1.3	<1.1	
Total-HxCDD	<1.4	<1.5	
Total-HpCDD	<2.3	<1.1	
Total-TCDF	<1.5	<1.4	
Total-PeCDF	<1.1	<0.82	
Total-HxCDF	1.11	<0.89	
Total-HpCDF	<1.4	<0.90	
<b>Toxic Equivalency - (WHO 2005)</b>			
<b>Lower Bound PCDD/F TEQ (WHO 2005)</b>	0.112	0.00341	
<b>Mid Point PCDD/F TEQ (WHO 2005)</b>	2.52	2.19	
<b>Upper Bound PCDD/F TEQ (WHO 2005)</b>	4.92	4.27	

# ALS Life sciences

## Continuing Calibration Summary Report

Sample Name	CCV	CCV
ALS Sample ID	H7-19-CCV-0534	H7-19-CCV-0535
Sample Size	1	1
Sample size units	n/a	n/a
Percent Moisture	n/a	n/a
Sample Matrix	QC	QC
Sampling Date	n/a	n/a
Extraction Date	n/a	n/a
<b>Target Analytes</b>	<b>% Rec</b>	<b>% Rec</b>
2,3,7,8-TCDD	122	121
1,2,3,7,8-PeCDD	114	110
1,2,3,4,7,8-HxCDD	96	91
1,2,3,6,7,8-HxCDD	111	104
1,2,3,7,8,9-HxCDD	99	95
1,2,3,4,6,7,8-HpCDD	110	113
OCDD	110	108
2,3,7,8-TCDF	125	125
1,2,3,7,8-PeCDF	109	112
2,3,4,7,8-PeCDF	125	114
1,2,3,4,7,8-HxCDF	98	88
1,2,3,6,7,8-HxCDF	114	109
2,3,4,6,7,8-HxCDF	108	111
1,2,3,7,8,9-HxCDF	104	110
1,2,3,4,6,7,8-HpCDF	103	104
1,2,3,4,7,8,9-HpCDF	105	104
OCDF	121	117
<b>Field Spike Standards</b>	<b>% Rec</b>	<b>% Rec</b>
37Cl4-2,3,7,8-TCDD	107	111
13C12-1,2,3,4,7,8-HxCDD	86	84
13C12-2,3,4,7,8-PeCDF	110	100
13C12-1,2,3,4,7,8-HxCDF	89	84
13C12-1,2,3,4,7,8,9-HpCDF	95	95
<b>Extraction Standards</b>		
13C12-2,3,7,8-TCDD	105	96
13C12-1,2,3,7,8-PeCDD	105	96
13C12-1,2,3,6,7,8-HxCDD	113	102
13C12-1,2,3,4,6,7,8-HpCDD	122	104
13C12-OCDD	111	112
13C12-2,3,7,8-TCDF	129	101
13C12-1,2,3,7,8-PeCDF	128	108
13C12-1,2,3,6,7,8-HxCDF	119	109
13C12-1,2,3,4,6,7,8-HpCDF	145	125
<b>Cleanup Standard</b>		
13C12-1,2,3,7,8,9-HxCDF	110	108

# **SVOC DATA PACKAGE**

## **SECTION 2b: INDIVIDUAL SAMPLE REPORTS**



# ALS Life Sciences

## Sample Analysis Report

**Sample Name** HEISER-25764410  
 ALS Sample ID L2310263-1  
 Analysis Method EPA T09A  
 Analysis Type Sample  
 Sample Matrix PUF

Sampling Date 11-Jul-19  
 Extraction Date 12-Aug-19  
 Sample Size 1 Puf  
 Percent Moisture n/a  
 Split Ratio 2

Approved:  
*T. Patterson*  
 --e-signature--  
 19-Aug-2019

**Run Information** **Run 1**  
 Filename 7-190816A13  
 Run Date 16-Aug-19 06:34  
 Final Volume 10 uL  
 Dilution Factor 1  
 Analysis Units pg  
 Instrument - Column HRMS-7 DB5MSUSR826231H

Target Analytes	TEF (WHO 2005)	Ret. Time	Conc. pg	EDL pg	Flags	EMPC pg	LQL
2,3,7,8-TCDD	1	NotFnd	<1.4	1.4	U		10
1,2,3,7,8-PeCDD	1	NotFnd	<0.61	0.61	U		50
1,2,3,4,7,8-HxCDD	0.1	NotFnd	<0.85	0.85	U		50
1,2,3,6,7,8-HxCDD	0.1	NotFnd	<0.78	0.78	U		50
1,2,3,7,8,9-HxCDD	0.1	NotFnd	<0.81	0.81	U		50
1,2,3,4,6,7,8-HpCDD	0.01	35.90	<1.3	0.75	M,J,R	1.3	50
OCDD	0.0003	37.40	<6.5	1.1	M,J,R	6.5	100
2,3,7,8-TCDF	0.1	NotFnd	<0.99	0.99	U		10
1,2,3,7,8-PeCDF	0.03	NotFnd	<0.65	0.65	U		50
2,3,4,7,8-PeCDF	0.3	NotFnd	<0.58	0.58	U		50
1,2,3,4,7,8-HxCDF	0.1	NotFnd	<0.65	0.65	U		50
1,2,3,6,7,8-HxCDF	0.1	NotFnd	<0.61	0.61	U		50
2,3,4,6,7,8-HxCDF	0.1	34.13	<0.65	0.65	M,U	0.47	50
1,2,3,7,8,9-HxCDF	0.1	34.53	<0.74	0.74	M,U		50
1,2,3,4,6,7,8-HpCDF	0.01	NotFnd	<0.55	0.55	U		50
1,2,3,4,7,8,9-HpCDF	0.01	NotFnd	<0.65	0.65	U		50
OCDF	0.0003	37.48	<1.5	0.83	M,J,R	1.5	100

Field Spike Standards	pg	% Rec	Limits
37Cl4-2,3,7,8-TCDD	600	28.14	82 70-130
13C12-1,2,3,4,7,8-HxCDD	6000	34.21	78 70-130
13C12-2,3,4,7,8-PeCDF	6000	31.94	79 70-130
13C12-1,2,3,4,7,8-HxCDF	6000	33.72	58 70-130
13C12-1,2,3,4,7,8,9-HpCDF	6000	36.14	69 70-130

Extraction Standards	pg	% Rec	Limits
13C12-2,3,7,8-TCDD	4000	28.11	70 40-130
13C12-1,2,3,7,8-PeCDD	4000	32.16	82 40-130
13C12-1,2,3,6,7,8-HxCDD	4000	34.26	84 40-130
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.89	81 25-130
13C12-OCDD	8000	37.39	68 25-130
13C12-2,3,7,8-TCDF	4000	27.19	86 40-130
13C12-1,2,3,7,8-PeCDF	4000	31.23	93 40-130
13C12-1,2,3,6,7,8-HxCDF	4000	33.79	88 40-130
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.33	90 25-130

Cleanup Standard	pg	% Rec	Limits
13C12-1,2,3,7,8,9-HxCDF	4000	34.54	86 40-130

Homologue Group Totals	# peaks	Conc. pg	EDL pg	LQL
Total-TCDD	0	<1.4	1.4	U 10
Total-PeCDD	0	<0.61	0.61	U 50
Total-HxCDD	0	<0.85	0.85	U 50
Total-HpCDD	0	<0.75	0.75	U 50
Total-TCDF	0	<0.99	0.99	U 10
Total-PeCDF	0	<0.65	0.65	U 50
Total-HxCDF	0	<0.74	0.74	U 50
Total-HpCDF	0	<0.65	0.65	U 50

Toxic Equivalency - (WHO 2005)	pg
Lower Bound PCDD/F TEQ (WHO 2005)	0.00
Mid Point PCDD/F TEQ (WHO 2005)	1.43
Upper Bound PCDD/F TEQ (WHO 2005)	2.84

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.  
 TEF Indicates the Toxic Equivalency Factor TEQ Indicates the Toxic Equivalency  
 M Indicates that a peak has been manually integrated.  
 U Indicates that this compound was not detected above the EDL.  
 J Indicates that a target analyte was detected below the calibrated range.  
 R Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.  
 LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.  
 EMPC Estimated Maximum Possible Concentration - elevated detection limit due to interference or positive id criterion failure

# ALS Life Sciences

## Sample Analysis Report

**Sample Name** CITY-45764407  
**ALS Sample ID** L2310263-2  
**Analysis Method** EPA T09A  
**Analysis Type** Sample  
**Sample Matrix** PUF

**Sampling Date** 11-Jul-19  
**Extraction Date** 12-Aug-19  
**Sample Size** 1 Puf  
**Percent Moisture** n/a  
**Split Ratio** 2

**Approved:**  
*T. Patterson*  
 --e-signature--  
 19-Aug-2019

**Run Information** **Run 1**  
**Filename** 7-190816A14  
**Run Date** 16-Aug-19 07:16  
**Final Volume** 10 uL  
**Dilution Factor** 1  
**Analysis Units** pg  
**Instrument - Column** HRMS-7 DB5MSUSR826231H

Target Analytes	TEF (WHO 2005)	Ret. Time	Conc. pg	EDL pg	Flags	EMPC pg	LQL
2,3,7,8-TCDD	1	NotFnd	<2.1	2.1	U		10
1,2,3,7,8-PeCDD	1	NotFnd	<1.0	1.0	U		50
1,2,3,4,7,8-HxCDD	0.1	NotFnd	<1.2	1.2	U		50
1,2,3,6,7,8-HxCDD	0.1	NotFnd	<1.1	1.1	U		50
1,2,3,7,8,9-HxCDD	0.1	NotFnd	<1.1	1.1	U		50
1,2,3,4,6,7,8-HpCDD	0.01	35.88	<1.4	0.68	M,J,R	1.4	50
OCDD	0.0003	37.38	8.24	0.85	M,J		100
2,3,7,8-TCDF	0.1	NotFnd	<1.2	1.2	U		10
1,2,3,7,8-PeCDF	0.03	NotFnd	<0.82	0.82	U		50
2,3,4,7,8-PeCDF	0.3	NotFnd	<0.73	0.73	U		50
1,2,3,4,7,8-HxCDF	0.1	NotFnd	<0.67	0.67	U		50
1,2,3,6,7,8-HxCDF	0.1	NotFnd	<0.62	0.62	U		50
2,3,4,6,7,8-HxCDF	0.1	NotFnd	<0.66	0.66	U		50
1,2,3,7,8,9-HxCDF	0.1	34.53	<1.1	0.76	M,J,R	1.1	50
1,2,3,4,6,7,8-HpCDF	0.01	35.31	<0.81	0.71	M,J,R	0.81	50
1,2,3,4,7,8,9-HpCDF	0.01	NotFnd	<0.84	0.84	U		50
OCDF	0.0003	37.47	<1.3	0.88	M,J,R	1.3	100

Field Spike Standards	pg	% Rec	Limits
37Cl4-2,3,7,8-TCDD	600	28.15	80 70-130
13C12-1,2,3,4,7,8-HxCDD	6000	34.20	78 70-130
13C12-2,3,4,7,8-PeCDF	6000	31.94	80 70-130
13C12-1,2,3,4,7,8-HxCDF	6000	33.71	56 70-130
13C12-1,2,3,4,7,8,9-HpCDF	6000	36.12	61 70-130

Extraction Standards	pg	Conc. pg	EDL pg
13C12-2,3,7,8-TCDD	4000	28.14	92 40-130
13C12-1,2,3,7,8-PeCDD	4000	32.15	119 40-130
13C12-1,2,3,6,7,8-HxCDD	4000	34.25	108 40-130
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.88	120 25-130
13C12-OCDD	8000	37.38	118 25-130
13C12-2,3,7,8-TCDF	4000	27.22	102 40-130
13C12-1,2,3,7,8-PeCDF	4000	31.23	120 40-130
13C12-1,2,3,6,7,8-HxCDF	4000	33.78	110 40-130
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.31	118 25-130

Cleanup Standard	pg	Conc. pg	EDL pg
13C12-1,2,3,7,8,9-HxCDF	4000	34.52	117 40-130

Homologue Group Totals	# peaks	Conc. pg	EDL pg		
Total-TCDD	0	<2.1	2.1	U	10
Total-PeCDD	0	<1.0	1.0	U	50
Total-HxCDD	0	<1.2	1.2	U	50
Total-HpCDD	0	<0.68	0.68	U	50
Total-TCDF	0	<1.2	1.2	U	10
Total-PeCDF	0	<0.82	0.82	U	50
Total-HxCDF	0	<0.76	0.76	U	50
Total-HpCDF	0	<0.84	0.84	U	50

Toxic Equivalency - (WHO 2005)	pg
Lower Bound PCDD/F TEQ (WHO 2005)	0.00247
Mid Point PCDD/F TEQ (WHO 2005)	2.14
Upper Bound PCDD/F TEQ (WHO 2005)	4.14

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.  
 TEF Indicates the Toxic Equivalency Factor TEQ Indicates the Toxic Equivalency  
 M Indicates that a peak has been manually integrated.  
 U Indicates that this compound was not detected above the EDL.  
  
 J Indicates that a target analyte was detected below the calibrated range.  
 R Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.  
  
 LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.  
 EMPC Estimated Maximum Possible Concentration – elevated detection limit due to interference or positive id criterion failure

# ALS Life Sciences

## Sample Analysis Report

**Sample Name** RES-25764409  
**ALS Sample ID** L2310263-3  
**Analysis Method** EPA T09A  
**Analysis Type** Sample  
**Sample Matrix** PUF

**Sampling Date** 11-Jul-19  
**Extraction Date** 12-Aug-19  
**Sample Size** 1 Puf  
**Percent Moisture** n/a  
**Split Ratio** 2

**Approved:**  
*T. Patterson*  
 --e-signature--  
 19-Aug-2019

**Run Information** **Run 1**  
**Filename** 7-190816A15  
**Run Date** 16-Aug-19 07:58  
**Final Volume** 10 uL  
**Dilution Factor** 1  
**Analysis Units** pg  
**Instrument - Column** HRMS-7 DB5MSUSR826231H

Target Analytes	TEF (WHO 2005)	Ret. Time	Conc. pg	EDL pg	Flags	EMPC pg	LQL
2,3,7,8-TCDD	1	NotFnd	<2.0	2.0	U		10
1,2,3,7,8-PeCDD	1	NotFnd	<0.89	0.89	U		50
1,2,3,4,7,8-HxCDD	0.1	NotFnd	<1.1	1.1	U		50
1,2,3,6,7,8-HxCDD	0.1	NotFnd	<1.0	1.0	U		50
1,2,3,7,8,9-HxCDD	0.1	NotFnd	<1.1	1.1	U		50
1,2,3,4,6,7,8-HpCDD	0.01	NotFnd	<2.4	2.4	U		50
OCDD	0.0003	37.39	<5.9	2.3	M,J,R	5.9	100
2,3,7,8-TCDF	0.1	NotFnd	<1.2	1.2	U		10
1,2,3,7,8-PeCDF	0.03	NotFnd	<0.93	0.93	U		50
2,3,4,7,8-PeCDF	0.3	NotFnd	<0.84	0.84	U		50
1,2,3,4,7,8-HxCDF	0.1	NotFnd	<0.70	0.70	U		50
1,2,3,6,7,8-HxCDF	0.1	NotFnd	<0.65	0.65	U		50
2,3,4,6,7,8-HxCDF	0.1	NotFnd	<0.69	0.69	U		50
1,2,3,7,8,9-HxCDF	0.1	34.54	<1.3	0.79	M,J,R	1.3	50
1,2,3,4,6,7,8-HpCDF	0.01	NotFnd	<0.89	0.89	U		50
1,2,3,4,7,8,9-HpCDF	0.01	NotFnd	<1.1	1.1	U		50
OCDF	0.0003	NotFnd	<1.7	1.7	U		100

Field Spike Standards	pg	% Rec	Limits
37Cl4-2,3,7,8-TCDD	600	28.14	87 70-130
13C12-1,2,3,4,7,8-HxCDD	6000	34.20	80 70-130
13C12-2,3,4,7,8-PeCDF	6000	31.94	91 70-130
13C12-1,2,3,4,7,8-HxCDF	6000	33.71	70 70-130
13C12-1,2,3,4,7,8,9-HpCDF	6000	36.13	75 70-130

Extraction Standards	pg	Conc. pg	EDL pg
13C12-2,3,7,8-TCDD	4000	28.11	64 40-130
13C12-1,2,3,7,8-PeCDD	4000	32.15	81 40-130
13C12-1,2,3,6,7,8-HxCDD	4000	34.25	82 40-130
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.88	76 25-130
13C12-OCDD	8000	37.38	57 25-130
13C12-2,3,7,8-TCDF	4000	27.19	85 40-130
13C12-1,2,3,7,8-PeCDF	4000	31.23	89 40-130
13C12-1,2,3,6,7,8-HxCDF	4000	33.78	86 40-130
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.32	83 25-130

Cleanup Standard	pg	Conc. pg	EDL pg
13C12-1,2,3,7,8,9-HxCDF	4000	34.52	89 40-130

Homologue Group Totals	# peaks	Conc. pg	EDL pg
Total-TCDD	0	<2.0	2.0 U 10
Total-PeCDD	0	<0.89	0.89 U 50
Total-HxCDD	0	<1.1	1.1 U 50
Total-HpCDD	0	<2.4	2.4 U 50
Total-TCDF	0	<1.2	1.2 U 10
Total-PeCDF	0	<0.93	0.93 U 50
Total-HxCDF	0	<0.79	0.79 U 50
Total-HpCDF	0	<1.1	1.1 U 50

Toxic Equivalency - (WHO 2005)	pg
Lower Bound PCDD/F TEQ (WHO 2005)	0.00
Mid Point PCDD/F TEQ (WHO 2005)	2.06
Upper Bound PCDD/F TEQ (WHO 2005)	3.99

EDL Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.  
 TEF Indicates the Toxic Equivalency Factor TEQ Indicates the Toxic Equivalency  
 M Indicates that a peak has been manually integrated.  
 U Indicates that this compound was not detected above the EDL.  
 J Indicates that a target analyte was detected below the calibrated range.  
 R Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.  
 LQL Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.  
 EMPC Estimated Maximum Possible Concentration - elevated detection limit due to interference or positive id criterion failure

# **SVOC DATA PACKAGE**

## **SECTION 3: METHOD SUMMARY**

**PCDD/F METHOD SUMMARY**  
**Methods 23/0023A/1613B/8290/TO-9A**

**Introduction:**

This summary is to provide ALSE Burlington PCDD/F method details in order to provide persons reviewing or validating this data package sufficient information to re-construct the sample calculation, data verification and review. It incorporates the analysis of PCDD/F via the following reference methods:

- US EPA Office of Water, Method 1613B
- US EPA Office of Solid Waste, SW846 Methods 8290A and 0023/8290A
- US EPA Office of Research & Development Method TO-9A.
- US EPA Office of Air Quality Planning & Standards Method 23.

Any deviations to what is listed herein would be listed in the project narrative.

To avoid the confusion and conflicting nomenclature within the methods, we have defined the labeled standards in terms relating to the time of addition to the sample or extract. Therefore;

- The Field or Sampling Standards are added prior to field sampling
- The Extraction Standards are added prior to extraction
- The Clean-up Standards are added prior to extract clean-up
- The Injection Standards are added prior to extract injection.

**Calibration Standard Levels:**

Six levels of standard are available for calibration as listed in Table 1. The low point (the CS0) is below method requirements and therefore is optional.

**Table 1: Calibration Standards**

	CS0	CS1	CS2	CS3	CS4	CS5	
<b>Natives</b>	2,3,7,8-TCDD	0.1	0.5	2	10	40	200
	2,3,7,8-TCDF	0.1	0.5	2	10	40	200
	1,2,3,7,8-PeCDD	0.5	2.5	10	50	200	1000
	1,2,3,7,8-PeCDF	0.5	2.5	10	50	200	1000
	2,3,4,7,8-PeCDF	0.5	2.5	10	50	200	1000
	1,2,3,4,7,8-HxCDD	0.5	2.5	10	50	200	1000
	1,2,3,6,7,8-HxCDD	0.5	2.5	10	50	200	1000
	1,2,3,7,8,9-HxCDD	0.5	2.5	10	50	200	1000
	1,2,3,4,7,8-HxCDF	0.5	2.5	10	50	200	1000
	1,2,3,6,7,8-HxCDF	0.5	2.5	10	50	200	1000
	1,2,3,7,8,9-HxCDF	0.5	2.5	10	50	200	1000
	2,3,4,6,7,8-HxCDF	0.5	2.5	10	50	200	1000
	1,2,3,4,6,7,8-HpCDD	0.5	2.5	10	50	200	1000
	1,2,3,4,6,7,8-HpCDF	0.5	2.5	10	50	200	1000
	1,2,3,4,7,8,9-HpCDF	0.5	2.5	10	50	200	1000
	OCDD	1	5	20	100	400	2000
	OCDF	1	5	20	100	400	2000
<b>Labeled</b>	2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	OCDD- <sup>13</sup> C <sub>12</sub>	200	200	200	200	200	200
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	0.1	0.5	2	10	40	200	
<b>Injection</b>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100
	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	100	100	100	100	100	100

## Calibration Control Limits

The initial and continuing calibration control limits for all methods are presented in Table 2 below. For the initial calibration CS1 and for each calibration verification CS3, the signal to noise ratio for each quantification ion for labelled and non-labelled analytes must be greater than or equal to 10:1

Table 2: Calibration Control Limits

	1613B		TO9A & TO-9A		8290A	
	Initial Cal.	Cal. Ver.	Initial Cal.	Cal. Ver.	Initial Cal.	Cal. Ver.
	%RSD	ng/mL	%RSD	% Diff	%RSD	% Diff
<b>Natives</b>						
2,3,7,8-TCDD	20	7.8-12.9	25	25	20	20*
2,3,7,8-TCDF	20	8.4-12.0	25	25	20	20*
1,2,3,7,8-PeCDD	20	39-65	25	25	20	20*
1,2,3,7,8-PeCDF	20	41-60	25	25	20	20*
2,3,4,7,8-PeCDF	20	41-61	25	25	20	20*
1,2,3,4,7,8-HxCDD	20	39-64	25	25	20	20*
1,2,3,6,7,8-HxCDD	20	39-64	25	25	20	20*
1,2,3,7,8,9-HxCDD	35	41-61	25	25	20	20*
1,2,3,4,7,8-HxCDF	20	45-56	25	25	20	20*
1,2,3,6,7,8-HxCDF	20	44-57	25	25	20	20*
1,2,3,7,8,9-HxCDF	20	45-56	25	25	20	20*
2,3,4,6,7,8-HxCDF	20	44-57	25	25	20	20*
1,2,3,4,6,7,8-HpCDD	20	43-58	25	25	20	20*
1,2,3,4,6,7,8-HpCDF	20	45-55	25	25	20	20*
1,2,3,4,7,8,9-HpCDF	20	43-58	25	25	20	20*
OCDD	20	79-126	25	25	20	20*
OCDF	35	63-159	30	30	20	20*
<b>Labels</b>						
2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	35	82-121	25	25	30	30**
2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	35	71-140	30	30	30	30**
1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	35	62-160	30	30	30	30**
1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	35	76-130	30	30	30	30**
2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	35	77-130	25	25	30	30**
1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	35	85-117	25	25	30	30**
1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	35	85-118	25	25	30	30**
1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	35	76-131	25	25	30	30**
1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	35	70-143	30	30	30	30**
1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	35	74-135	-	-	-	-
2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	35	73-137	30	30	30	30**
1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	35	72-138	30	30	30	30**
1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	35	78-129	30	30	30	30**
1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	35	77-129	25	25	30	30**
OCDD- <sup>13</sup> C <sub>12</sub>	35	96-415	30	30	30	30**
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	35	7.9-12.7	25	25	30	30**

\* 25% is allowed for a post-run verification but when the value is above 20%, then the analyte quantification must be as per 8290A Section 8.3.2.4 and corrective action is required before more samples can be analyzed.

\*\*35% is allowed for a post-run verification but when the value is above 30%, then the analyte quantification must be as per 8290A Section 8.3.2.4 and corrective action is required before more samples can be analyzed.

**LCS Criteria:**

The laboratory control sample (LCS) or the On-Going Precision and Accuracy (OPR) recovery criteria are listed in Table 3

**Table 3: Acceptance Criteria for IPR and OPR<sup>a</sup>**

	Test Conc.	IPR		OPR
		s <sup>b</sup>	X <sup>c</sup>	
	ng/L	ng/L	ng/L	ng/L
<b>Natives</b>				
2,3,7,8-TCDD	10	2.8	8.3-12.9	6.7-15.8
2,3,7,8-TCDF	10	2	8.7-13.7	7.5-15.8
1,2,3,7,8-PeCDD	50	7.5	38-66	35-71
1,2,3,7,8-PeCDF	50	7.5	43-62	40-67
2,3,4,7,8-PeCDF	50	8.6	36-75	34-80
1,2,3,4,7,8-HxCDD	50	9.4	39-76	35-82
1,2,3,6,7,8-HxCDD	50	7.7	42-62	38-67
1,2,3,7,8,9-HxCDD	50	11.1	37-71	32-81
1,2,3,4,7,8-HxCDF	50	8.7	41-59	36-67
1,2,3,6,7,8-HxCDF	50	6.7	46-60	42-65
1,2,3,7,8,9-HxCDF	50	6.4	42-61	39-65
2,3,4,6,7,8-HxCDF	50	7.4	37-74	35-78
1,2,3,4,6,7,8-HpCDD	50	7.7	38-65	35-70
1,2,3,4,6,7,8-HpCDF	50	6.3	45-56	41-61
1,2,3,4,7,8,9-HpCDF	50	8.1	43-63	39-69
OCDD	100	19	89-127	78-144
OCDF	100	27	74-146	63-170
<b>Labels</b>				
2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	100	37	28-134	20-175
2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	100	35	31-113	22-152
1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	100	39	27-184	21-227
1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	100	34	27-156	21-192
2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	100	38	16-297	13-328
1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	100	41	29-147	21-193
1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	100	38	34-122	25-163
1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	43	27-152	19-202
1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	35	30-122	21-159
1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	100	40	24-157	17-205
2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	100	37	29-136	22-176
1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	100	35	34-129	26-166
1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	100	41	32-110	21-158
1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	100	40	28-141	20-186
OCDD- <sup>13</sup> C <sub>12</sub>	200	95	41-276	26-397
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	10	3.6	3.9-15.4	3.1-19.1

<sup>a</sup> Assuming a final volume of 20uL

<sup>b</sup> s = standard deviation

<sup>c</sup> X = Average Concentration

## Extraction/Clean-up & Sampling Standard Recovery Limits:

Table 4: Extraction, Clean-up, Injection & Sampling Standard Recovery Limits

	1613B or 8290A (non Stack)		TO9A or 0023A/8290A or TO-9A	
	(% Rec.)	Ref.	(% Rec.)	Ref.
<b>Extraction Standard</b>				
2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	25-164	a	40-130	b
2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	24-169	a	40-130	b
1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	25-181	a	40-130	b
1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	24-185	a	40-130	b
2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	21-178	a	-	
1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	32-141	a	-	
1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	28-130	a	40-130	b
1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	26-152	a	-	
1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	26-123	a	40-130	b
1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	29-147	a	-	
2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	28-136	a	40-130	c,d
1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	23-140	a	25-130	b
1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	28-143	a	25-130	b
1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	26-138	a	-	
OCDD- <sup>13</sup> C <sub>12</sub>	17-157	a	25-130	b
<b>Clean-up Standard</b>				
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	35-197	a	-	
1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	-		40-130	b
<b>Injection Standard</b>				
1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	30-300	d	30-300	d
1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	30-300	d	30-300	d
<b>Sampling Standard</b>				
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	-		70-130	b
2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	-		70-130	b
1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	-		70-130	b
1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	-		70-130	b
1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	-		70-130	b

### References & Notes

<sup>a</sup> from OW method 1613B

<sup>b</sup> from OAQPS method 23

<sup>c</sup> this extraction standard is not required in methods 23 and 0023A/8290A

<sup>d</sup> ALS In-house criteria



### **Reporting Limits:**

Unless indicated in the otherwise, the PCDD/F data is reported down to 2.5:1 signal to noise for each isomer grouping for each extract injection. This is consistent to SW846 8290 defined protocols (i.e. EDL or Estimated Detection Limit) and is commonly applied throughout the industry to all the HRMS PCDD/F methods applicable to this method summary.

### **Method Blank:**

The method blank levels must be below the response to the low calibration standard, CS0 or CS1, whichever low calibration point is being applied to the project.

### **MS/MSD:**

The % relative difference between the MS and MSD spike recoveries should be less than or equal to 20%.

### **Instrument/Run Performance Criteria:**

- 1 Elution windows must be defined by a 'Window Performance Mix' at the beginning of each 12-hour run sequence
- 2 GC performance criteria of 25% maximum valley between 2,3,7,8-TCDD and it's neared eluting isomers (DB5) or 2,3,7,8-TCDF and it's nearest eluting isomers (DB225).
- 3 At the beginning of and just following the end of each 12 hour run sequence, the instrument must be checked to demonstrate a resolution of 10,000 for each quantification window.
- 4 For method 1613B, the relative retention times (RRT) of the compounds in the daily CS3 calibration verification must fall into the ranges presented in Table 4.
- 5 For all calibrations, QC samples and field samples, the absolute retention time (RT) for 1,2,3,4-TCDD-13C12 must be >25.0 min on a DB5 column and >15.0 min on a DB225 column.
- 6 The RT in the daily CS3 verification standards must be within 15 seconds of the CS3 in the initial calibration run.
- 7 The maximum time between scans within a descriptor is 1 second.
- 8 Lock mass deviations to the average response must be less than or equal 20%.

### **Laboratory Duplicates:**

The % relative difference between duplicates should be less than or equal to 25% but only where the response is greater than the low calibration standard.

### **Analyte Identification Criteria:**

- 1 Ion ratio must be within 15% of theoretical or within 10% of the most recent CS3.
- 2 The retention time (RT) of the peak maxima for each pair of quantification ions must be no more than 2 seconds (i.e. 2 scans) difference.
- 3 The retention time (RT) of the peak maxima of all 2,3,7,8- substituted native analytes must be within -1 to +3 seconds of the RT of corresponding <sup>13</sup>C<sub>12</sub>-labelled isomer of that injection run.
- 4 For those native analytes without a corresponding labelled isomer, the relative retention time (RRT) must be within 0.005 of the relative retention time observed in the daily CS3 run.
- 5 When there is a significant PCDPE interference observed, then a peak in the PCDF channel is not confirmed to be PCDF. [Significant PCDPE interference is identified when there is a PCDPE parent ion peak 10% or more of the response of a peak at the same RT (i.e. within 2 seconds) in the corresponding PCDF channel. ]
- 6 For any peak to be identified as a positive PCDD/F response, that peak must be within the retention time windows defined by the daily analysis of Window Performance Mixture.

**Table 4: Quantitation References and Method 1613B RT References and RRT**

Analyte	Stack/Ambient Quantitation Reference	Method 1613B RT Reference	Method 1613B RRT
		Solids/ Waters Quantitation Reference	
<b>Compounds using 1,2,3,4-TCDD-<sup>13</sup>C<sub>12</sub> as injection standard</b>			
2,3,7,8-TCDF	2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.003
2,3,7,8-TCDD	2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.002
1,2,3,7,8-PeCDF	1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.002
2,3,4,7,8-PeCDF	1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.002
1,2,3,7,8-PeCDD	1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.002
2,3,7,8-TCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	0.923-1.103
2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	0.976-1.043
2,3,7,8-TCDD- <sup>37</sup> Cl <sub>4</sub>	2,3,7,8-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	0.989-1.052
1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1.000-1.425
2,3,4,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8-PeCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1.011-1.526
1,2,3,7,8-PeCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4-TCDD- <sup>13</sup> C <sub>12</sub>	1.000-1.567
<b>Compounds using 1,2,3,7,8,9-HxCDD-<sup>13</sup>C<sub>12</sub> as injection standard</b>			
1,2,3,4,7,8-HxCDF	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,6,7,8-HxCDF	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	0.997-1.005
1,2,3,7,8,9-HxCDF	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.001
2,3,4,6,7,8-HxCDF	2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,4,7,8-HxCDD	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,6,7,8-HxCDD	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	0.998-1.004
1,2,3,7,8,9-HxCDD <sup>a</sup>	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	<sup>a</sup>	1.000-1.019
1,2,3,4,6,7,8-HpCDF	1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,4,7,8,9-HpCDF	1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,4,6,7,8-HpCDD	1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.001
OCDF	OCDD- <sup>13</sup> C <sub>12</sub>	OCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.008
OCDD	OCDD- <sup>13</sup> C <sub>12</sub>	OCDD- <sup>13</sup> C <sub>12</sub>	0.999-1.001
1,2,3,4,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.944-0.970
1,2,3,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.949-0.975
1,2,3,7,8,9-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.977-1.047
2,3,4,6,7,8-HxCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.959-1.021
1,2,3,4,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.977-1.000
1,2,3,6,7,8-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	0.981-1.003
1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1.043-1.085
1,2,3,4,7,8,9-HpCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,4,6,7,8-HpCDF- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1.057-1.151
1,2,3,4,6,7,8-HpCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1.086-1.110
OCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1,2,3,7,8,9-HxCDD- <sup>13</sup> C <sub>12</sub>	1.032-1.311

<sup>a</sup> For solids/waters via 1612B, 1,2,3,7,8,9-HxCDD is quantified against the average responses of 1,2,3,4,7,8-HxCDD-<sup>13</sup>C<sub>12</sub> and 1,2,3,6,7,8-HxCDD-<sup>13</sup>C<sub>12</sub> while 1,2,3,6,7,8-HxCDD-<sup>13</sup>C<sub>12</sub> is the RT reference.

**Table 5: HRMS Instrumental Descriptor Parameters**

Descriptor	Exact M/Z	M/Z Type	Elemental Composition	Substance	Type	Theoretical	Ion Ratio QC Limits		
						Ion Ratio	Upper	Lower	
1	303.9016	M	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>4</sub> <sup>16</sup> O	TCDF	Native	0.77	0.65	0.89	
	305.8987	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sup>16</sup> O	TCDF	Native				
	315.9419	M	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>4</sub> <sup>16</sup> O	TCDF	<sup>13</sup> C	0.77	0.65	0.89	
	317.9389	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sup>16</sup> O	TCDF	<sup>13</sup> C				
	316.9824	Lock	<sup>12</sup> C <sub>9</sub> <sup>19</sup> F <sub>11</sub>	PFK	Lock				
	319.8965	M	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>4</sub> <sup>16</sup> O <sub>2</sub>	TCDD	Native	0.77	0.65	0.89	
	321.8936	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	TCDD	Native				
	327.8847	M+8	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>37</sup> Cl <sub>4</sub> <sup>16</sup> O <sub>2</sub>	TCDD	<sup>37</sup> Cl				
	331.9368	M	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>4</sub> <sup>16</sup> O <sub>2</sub>	TCDD	<sup>13</sup> C	0.77	0.65	0.89	
	333.9339	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	TCDD	<sup>13</sup> C				
	339.8597	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O	PeCDF	Native	1.55	1.32	1.78	
	341.8568	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	PeCDF	Native				
	351.9	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O	PeCDF	<sup>13</sup> C	1.55	1.32	1.78	
	353.897	M+4	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	PeCDF	<sup>13</sup> C				
	375.8364	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>4</sub> <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sup>16</sup> O	HxCDFPE	CI-DPE				
	409.7974	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sup>16</sup> O	HpCDFPE	CI-DPE				
	2	339.8597	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O	PeCDF	Native	1.55	1.32	1.78
		341.8568	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	PeCDF	Native			
		351.9	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O	PeCDF	<sup>13</sup> C	1.55	1.32	1.78
		353.897	M+4	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>3</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	PeCDF	<sup>13</sup> C			
353.8576		M	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>5</sub> <sup>16</sup> O <sub>2</sub>	PeCDD	Native	0.63	0.54	0.72	
355.8546		M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	PeCDD	Native				
366.9792		Lock	<sup>12</sup> C <sub>10</sub> <sup>19</sup> F <sub>13</sub>	PFK	Lock				
365.8978		M	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>5</sub> <sup>16</sup> O <sub>2</sub>	PeCDD	<sup>13</sup> C	0.63	0.54	0.72	
367.8949		M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	PeCDD	<sup>13</sup> C				
409.7974		M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>3</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sup>16</sup> O	HpCDFPE	CI-DPE				
3		373.8207	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sup>16</sup> O	HxCDF	Native	1.24	1.05	1.43
		375.8178	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	HxCDF	Native			
		380.976	Lock	<sup>12</sup> C <sub>8</sub> <sup>19</sup> F <sub>5</sub>	PFK	Lock			
	383.8639	M	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>6</sub> <sup>16</sup> O	HxCDF	<sup>13</sup> C	0.51	0.43	0.59	
	385.861	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sup>16</sup> O	HxCDF	<sup>13</sup> C				
	389.8156	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	HxCDD	Native	1.24	1.05	1.43	
	391.8127	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	HxCDD	Native				
	401.8559	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	HxCDD	<sup>13</sup> C	1.24	1.05	1.43	
	403.853	M+4	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	HxCDD	<sup>13</sup> C				
	445.7555	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sub>2</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	OCDFPE	CI-DPE				
4	409.7789	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	HpCDF	Native	1.88	1.60	2.16	
	411.7759	M+6	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>4</sub> <sup>37</sup> Cl <sub>3</sub> <sup>16</sup> O	HpCDF	Native				
	417.8253	M	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>7</sub> <sup>16</sup> O	HpCDF	<sup>13</sup> C	0.44	0.37	0.51	
	419.822	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sup>16</sup> O	HpCDF	<sup>13</sup> C				
	423.7767	M+2	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	HpCDD	Native	1.04	0.88	1.20	
	425.7737	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	HpCDD	Native				
	430.9728	Lock	<sup>12</sup> C <sub>9</sub> <sup>19</sup> F <sub>17</sub>	PFK	Lock				
	435.8169	M+2	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	HpCDD	<sup>13</sup> C	1.04	0.88	1.20	
	437.814	M+4	<sup>13</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>5</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	HpCDD	<sup>13</sup> C				
	479.7165	M+4	<sup>12</sup> C <sub>12</sub> <sup>1</sup> H <sup>35</sup> Cl <sub>7</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	NCDPE	CI-DPE				
5	441.7428	M+2	<sup>12</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>7</sub> <sup>37</sup> Cl <sup>16</sup> O	OCDF	Native	0.89	0.76	1.02	
	443.7399	M+4	<sup>12</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	OCDF	Native				
	454.9728	Lock	<sup>12</sup> C <sub>11</sub> <sup>19</sup> F <sub>17</sub>	PFK	Lock				
	457.7377	M+2	<sup>12</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>7</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	OCDD	Native	0.89	0.76	1.02	
	459.7348	M+4	<sup>12</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	OCDD	Native				
	469.778	M+2	<sup>13</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>7</sub> <sup>37</sup> Cl <sup>16</sup> O <sub>2</sub>	OCDD	<sup>13</sup> C	0.89	0.76	1.02	
	471.775	M+4	<sup>13</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>6</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O <sub>2</sub>	OCDD	<sup>13</sup> C				
	513.6775	M+4	<sup>12</sup> C <sub>12</sub> <sup>35</sup> Cl <sub>8</sub> <sup>37</sup> Cl <sub>2</sub> <sup>16</sup> O	DCDFPE	CI-DPE				

## Data Calculations:

### a) Analyte Concentrations:

The relative response factor of each target relative to the standard against which it is to be calculated is determined using the area responses of both quantification ions via equation 9.1.

In cases where a native target is calculated against an exact labelled analogue, the quantification will be considered to be by isotope dilution. In other cases, the quantification will be considered to be by internal standard.

$$\text{RRF} = \frac{(A1_t + A2_t) C_s}{(A1_s + A2_s) C_t} \quad \text{Equ. 9.1}$$

Where,

$A1_t + A2_t$  = The areas of the two quantification ions for the target analyte

$A1_s + A2_s$  = The areas of the two quantification ions for the labelled compound against which the target analyte will be calculated.

$C_t$  = The concentration in the calibration standard of the target analyte.

$C_s$  = The concentration in the calibration standard of the labelled compound against which the target will be calculated.

For all analytes to be quantified and from the initial calibration series of standard injections, a table of RRFs is prepared. The relative standard deviation (%RSD, or the coefficient of variance) is checked to confirm that the appropriate method criteria has been met as listed in Table 3. The average of the five or six levels of standard for each analyte,  $\text{RRF}_{av}$  is applied for quantification of samples according to Equations 9.2 and 9.3 below.

$$\text{Amount in sample (pg)} = \frac{(A1_n + A2_n) Q_i}{(A1_i + A2_i) (\text{RRF}_{av})} \quad \text{Equ. 9.2}$$

$$\text{Concentration in sample (pg/g or pg/l)} = \frac{(A1_n + A2_n) Q_i}{(A1_i + A2_i) (\text{RRF}_{av}) (W_s)} \quad \text{Equ. 9.3}$$

Where,

$Q_i$  = The amount (pg) of labelled compound added to the sample

$W_s$  = The weight (g) or volume (l) of sample

### b) Extraction, Clean-up, and Sampling Standard Recovery Calculation:

The extraction, clean-up, and sampling standard recoveries are determined by Equation 9.4 below.

$$\% \text{ Recovery} = \frac{\text{Amount in sample}}{\text{Amount added to sample}} \times 100 \quad \text{Equ. 9.4}$$

### c) Estimated Detection Limit

$$\text{EDL} = \frac{2.5 \times H_x \times Q_{\text{es}}}{H_{\text{es}} \times W \times \text{RRF}_{\text{av}}} \quad \text{Equ. 9.5}$$

Where,

EDL = estimated detection limit for homologous 2,3,7,8-Substituted PCDD/Fs

$H_x$  = sum of the height of the noise level for each quantification ions for the unlabelled PCDD/Fs.

$H_{\text{es}}$  = Sum of the heights of responses of both quantification ions for the labelled extraction standard.

$W$  = weight of volume of sample

$\text{RRF}_{\text{av}}$  = average relative response factor

$Q_{\text{es}}$  = Amount of extraction standard added

## Chromatogram Annotation Codes

All manually integrated peaks are expanded and reprinted with the following annotations:

* Analyst Initials	AA
* Date	YYMMDD
* integration code	CC

The Syntax is:

AAYYMMDDCC

Example:

SK111220MB

Code	Mnemonic	Description
MB	Manual Baseline	The peak was manually integrated because the initial baseline was determined incorrectly by the software
MS	Manual Split	The peak was manually integrated because the peak was incorrectly or not split by the software
MJ/MC	Manual Join/Manual Combine	The peak was manually integrated because the peak was split by the software and the peak should be integrated as a single peak
MA	Manual Add	The peak was manually integrated because the signal:noise ratio was judged to be >2.5
MD	Manual Delete	The peak was excluded because the signal:noise ratio was judged to be <2.5
MX	Manual Exclude	The peak was excluded due to an interference
NH	Noise Height	The noise height for detection limit calculation was manually defined, over-riding the software chosen value
MT	Manual Time	The peak retention time was manually chosen

The following explanatory annotation codes may appear on the chromatograms of peaks that have been reviewed:

Code	Mnemonic	Description
+	Detected Peak	A peak was detected at this mass and retention time that was above 2.5:1 signal to noise
<	Below Detection Limit	The signal at this mass and retention time was below 2.5:1 signal to noise
EMPC	Estimated Maximum Possible Concentration	The signal at this mass and retention time is an interference such that the target compound could not be confirmed
X-RT	Not Detected due to Retention Time non-conformance	The signal at this retention time could not be used to positively identify the target compound because of retention time non-conformance (apex of quantification and confirmation ions do not maximize within the same two seconds, or the retention time of the peak does not fall within the expected range with respect to its labeled analogue)
X-LOC	Not Detected due to interference from a higher level of chlorination	The signal at this retention time is attributable to a fragment from a co-eluting compound at a higher level of chlorination, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)
X-DPE	Not Detected due to diphenyl ether interference	The signal at this retention time is attributable to interference from a chlorinated diphenyl ether, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)
X-IF	Not Detected due to interference	The signal at this retention time is attributable to a co-eluting interference, and cannot be used to positively identify the target. The result is expressed as an Estimated Maximum Possible Concentration (EMPC)

## **Deviations from the Primary Reference Methods:**

The following changes and clarifications apply:

1) The calibration standards as listed in Table 2 are applied appropriately to all of the reference methods listed above. Such an application of one standard calibration series to all of these methods is within the scope of each and every one of the methods. The calibration standard set CS1 through CS5 is consistent with the standards concentration listing in method 1613B Table 4. The CS0 extends the calibration range below what is required by all of the methods. Table 4 defines the use of each of the labelled standards relative to each of the methods.

a. Method 1613B lists a larger suite of labelled extraction standards than does method 8290A. Additional labelled extraction standards have been added into the 8290A analysis to enhance the method and the data quality. These additions to the method constitute performance based enhancements and are within the scope of SW846 Method 8290A.

b. The levels presented in the calibration table of method 8290A are recommended values only. Changes to these concentrations, especially to expand the range, are within the scope of the method. Therefore application of the 1613B calibration standards to method 8290A is compliant with the scope of the method.

c. TO-9A is also a performance based method. It specifically states that different extraction standards and different concentrations of standards from those listed in TO-9A Table 3 is acceptable (see Section 6.8 of reference method).

d. Although OAQPS reference method 23 is not a performance based method, application of the 1613B standards has been defined as within the scope of the method. (see Appendix B)

2) Chlorinated Diphenyl Ether interferences: Both methods 1613B and 8290A indicate that any instrumental response showing the presence of a chlorinated diphenyl ether response and that coelutes with a PCDF represents an interference on that analyte (see Sections 18.3 and 7.8.4.4 respectively). This apparent zero tolerance does not take into account that the response in the diphenyl ether channel may be trivial relative to the corresponding PCDF. For this 'Standard Method', we have defined a chlorinated diphenyl ether interference as the presence of a **significant** response within the chlorinated diphenyl ether channel (rather than zero response) and defined significant as a response equal to or greater than 10% of the peak response in the PCDF channel.

3) When the primary analysis is performed using a DB5MS GC column, 2,3,7,8-TCDF can be resolved to a valley height of 60% from the closest-eluting isomers for this column, providing good quantification of this target without further confirmation. Confirmation of 2,3,7,8-TCDF concentrations above the level of the lowest calibration standard are performed on a second column on a contract basis when requested. Confirmation of additional 2,3,7,8-substituted PCDD/F isomers is also available when requested.

4) Although not categorically stated in all associated PCDD/F methods, we maintain that each and every individual clean-up procedure is, by definition, performance-based and optional. There is not an expectation within the industry to follow exactly the descriptions of clean-ups in reference methods. Adaptations which meet or exceed the required performance criteria are therefore acceptable within the scope of each reference method. The reference method descriptions are intended as guidelines or templates available to help the laboratory to define effective in-house clean-up methods. The objective within the laboratory is to provide quality clean extracts to the instrument for analysis. Each individual clean-up is part of the laboratory's 'arsenal' in order to achieve this objective.

5) There are differences within the individual reference methods as to the precise spiking protocols for adding extraction standards and native spikes (for LCS, MS and MSD). To ensure consistency within the laboratory between PCDD/F and related methods, the PCDD/F preparative 'Specific Method' requires solids (including stack and ambient sorbants/filters) to be spiked in the Soxhlet thimble from a nonane solution and waters are spiked before filtering from an acetone solution. This is consistent with the 8290A approach.

6) Sub-sampling of solids and pre-extraction processing is done in a manner that minimizes potential for cross-contamination. These processes are designed around SW846 protocols rather than 1613B protocols. Solids are sub-sampled directly from the bottle as submitted to the laboratory wherever practical. If the sample is submitted such that homogenization in the bottle is impractical (eg. the bottle is too full or lumps cannot be broken down), then transferring the sample to a tray or another bottle may be in order.

7) The concentrations of labelled and native spiking solutions are not consistent with those listed in all of the reference methods. These concentrations are prepared at levels convenient and expedient for accurate laboratory processing.

8) With respect to extraction standard recovery limits on non-stack samples analyzed via method 8290A, the limits are based upon the inter-laboratory performance limits defined in method 1613B rather than the relatively arbitrary limits of 35-140% suggested in Section 8.4 of method 8290A.

9) With respect to ions monitored for P5CDD and H7CDF:

a. The 358 ion has a potential for interference from PCB (hexachlorobiphenyls) dependent upon levels of PCBs in the sample and the instrument tuning. Of particular concern is PCB-169 which on a DB5MS column elutes very close to 1,2,3,7,8-P5CDD and which is not removed for the PCDD/F extracts even by carbon clean-up. To eliminate the potential of such interferences from PCB on the 358 mass, we choose to monitor the alternate ion pair of 354 and 356.

b. Similarly, the 408 ion of native H7CDF is prone to problematic interferences arising from 13C12-labeled heptachlorinated biphenyls. To eliminate the potential of such interferences from PCB on the 358 mass, we choose to monitor the alternate ion pair of 410 and 412.

## **SVOC DATA PACKAGE**

### **SECTION 4: CALIBRATION DATA**

Including:

for Multi-Point Calibration(s)

- Multi-Point Calibration Tables
- Individual Quantitation Reports

for Continuing Calibration(s)

- Individual Quantitation Reports



# ALS Life Sciences

## Calibration Summary Report

Calibration Level	Filename	Run Date
CS-1	7-190724A03	24-Jul-2019 17:20
CS-2	7-190724A02	24-Jul-2019 16:37
CS-3	7-190724A01	24-Jul-2019 15:56
CS-4	7-190724A07	24-Jul-2019 20:08
CS-5	7-190724A06	24-Jul-2019 19:26

Approved:	<i>T.Patterson</i> --e-signature-- 14-Aug-2019
-----------	--

Target Analytes	Relative Response Factors					Mean	% RSD	CVS % Rec
	CS-1	CS-2	CS-3	CS-4	CS-5			
<b>2,3,7,8-TCDD</b>	0.987	0.991	1.261	1.155	1.148	1.108	11%	104%
<b>1,2,3,7,8-PeCDD</b>	0.816	0.895	0.923	0.948	0.914	0.899	6%	110%
<b>1,2,3,4,7,8-HxCDD</b>	0.879	0.923	0.965	0.993	0.955	0.943	5%	100%
<b>1,2,3,6,7,8-HxCDD</b>	0.828	0.914	0.940	1.005	0.944	0.926	7%	98%
<b>1,2,3,7,8,9-HxCDD</b>	0.880	0.920	0.956	1.027	0.935	0.944	6%	106%
<b>1,2,3,4,6,7,8-HpCDD</b>	0.836	0.884	0.946	1.000	0.965	0.926	7%	100%
<b>OCDD</b>	0.927	0.937	0.938	1.010	0.962	0.955	4%	98%
<b>2,3,7,8-TCDF</b>	0.783	0.779	0.913	0.904	0.870	0.850	8%	98%
<b>1,2,3,7,8-PeCDF</b>	0.790	0.834	0.864	0.908	0.881	0.855	5%	107%
<b>2,3,4,7,8-PeCDF</b>	0.808	0.917	0.911	0.970	0.941	0.909	7%	95%
<b>1,2,3,4,7,8-HxCDF</b>	0.983	1.056	1.064	1.099	1.083	1.057	4%	100%
<b>1,2,3,6,7,8-HxCDF</b>	0.937	0.964	0.993	1.041	0.984	0.984	4%	106%
<b>2,3,4,6,7,8-HxCDF</b>	0.986	1.034	1.081	1.121	1.086	1.062	5%	99%
<b>1,2,3,7,8,9-HxCDF</b>	0.895	0.926	1.006	1.031	1.002	0.972	6%	108%
<b>1,2,3,4,6,7,8-HpCDF</b>	0.794	0.849	0.888	0.936	0.910	0.875	6%	99%
<b>1,2,3,4,7,8,9-HpCDF</b>	0.850	0.867	0.939	0.944	0.920	0.904	5%	103%
<b>OCDF</b>	1.005	1.081	1.082	1.185	1.137	1.098	6%	104%
<b>Extraction Standards</b>								
<b>13C12-2,3,7,8-TCDD</b>	0.897	0.899	0.909	0.908	0.934	0.909	2%	
<b>13C12-1,2,3,7,8-PeCDD</b>	0.688	0.726	0.725	0.748	0.842	0.746	8%	
<b>13C12-1,2,3,4,7,8-HxCDD</b>	0.910	1.010	0.992	0.986	1.025	0.985	5%	
<b>13C12-1,2,3,6,7,8-HxCDD</b>	1.088	1.109	1.099	1.083	1.094	1.095	1%	
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>	0.775	0.893	0.865	0.815	0.863	0.842	6%	
<b>13C12-OCDD</b>	0.641	0.776	0.766	0.729	0.857	0.754	10%	
<b>13C12-2,3,7,8-TCDF</b>	1.421	1.421	1.462	1.421	1.467	1.438	2%	
<b>13C12-1,2,3,7,8-PeCDF</b>	1.000	1.046	1.047	1.042	1.181	1.063	6%	
<b>13C12-2,3,4,7,8-PeCDF</b>	1.042	1.072	1.091	1.112	1.240	1.111	7%	
<b>13C12-1,2,3,4,7,8-HxCDF</b>	1.109	1.175	1.166	1.124	1.127	1.140	3%	
<b>13C12-1,2,3,6,7,8-HxCDF</b>	1.263	1.385	1.362	1.280	1.303	1.319	4%	
<b>13C12-2,3,4,6,7,8-HxCDF</b>	1.114	1.205	1.159	1.111	1.131	1.144	3%	
<b>13C12-1,2,3,7,8,9-HxCDF</b>	1.029	1.121	1.108	1.085	1.109	1.090	3%	
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>	0.885	1.008	0.957	0.909	0.934	0.939	5%	
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>	0.695	0.808	0.779	0.756	0.800	0.768	6%	
<b>Cleanup Standard</b>								
<b>37Cl4-2,3,7,8-TCDD (Cleanup)</b>	0.967	0.973	1.094	1.039	1.179	1.050	8%	

**CVS** Indicates a second source Calibration Verification Standard

# ALS Life Sciences

## Calibration Report

ALS Sample ID **H7-19-CCV-CS1-0017**  
 Analysis Method EPA TO9A  
 Analysis Type Calibration

Filename 7-190724A03      Inst # HRMS-7      Column DB5MSUSR826231H      Run Date 24-Jul-2019 17:20

Approved: *T.Patterson*  
 --e-signature--  
 19-Aug-2019

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
<b>2,3,7,8-TCDD</b>	28.05	0.71	0.50	5.56E+03	1.000
<b>1,2,3,7,8-PeCDD</b>	32.12	1.53	2.50	1.74E+04	0.816
<b>1,2,3,4,7,8-HxCDD</b>	34.16	1.22	2.50	1.57E+04	0.734
<b>1,2,3,6,7,8-HxCDD</b>	34.21	1.30	2.50	1.77E+04	0.828
<b>1,2,3,7,8,9-HxCDD</b>	34.34	1.21	2.50	1.72E+04	0.808
<b>1,2,3,4,6,7,8-HpCDD</b>	35.82	1.01	2.50	1.27E+04	0.836
<b>OCDD</b>	37.32	0.90	5.00	2.33E+04	0.927
<b>2,3,7,8-TCDF</b>	27.13	0.75	0.50	6.88E+03	0.781
<b>1,2,3,7,8-PeCDF</b>	31.18	1.55	2.50	2.45E+04	0.790
<b>2,3,4,7,8-PeCDF</b>	31.89	1.57	2.50	2.61E+04	0.841
<b>1,2,3,4,7,8-HxCDF</b>	33.67	1.13	2.50	2.14E+04	0.863
<b>1,2,3,6,7,8-HxCDF</b>	33.73	1.16	2.50	2.32E+04	0.937
<b>2,3,4,6,7,8-HxCDF</b>	34.06	1.16	2.50	2.15E+04	0.869
<b>1,2,3,7,8,9-HxCDF</b>	34.48	1.23	2.50	1.81E+04	0.729
<b>1,2,3,4,6,7,8-HpCDF</b>	35.27	1.90	2.50	1.38E+04	0.794
<b>1,2,3,4,7,8,9-HpCDF</b>	36.08	1.79	2.50	1.16E+04	0.667
<b>OCDF</b>	37.41	0.89	5.00	2.53E+04	1.005
<b>Field Spike Standards</b>					
<b>37Cl4-2,3,7,8-TCDD</b>	28.06	0.00	0.50	6.00E+03	1.079
<b>13C12-1,2,3,4,7,8-HxCDD</b>	34.15	1.26	100.00	7.13E+05	0.836
<b>13C12-2,3,4,7,8-PeCDF</b>	31.88	1.56	100.00	1.29E+06	1.042
<b>13C12-1,2,3,4,7,8-HxCDF</b>	33.66	0.51	100.00	8.70E+05	0.878
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>	36.07	0.46	100.00	5.45E+05	0.785
<b>Extraction Standards</b>					
<b>13C12-2,3,7,8-TCDD</b>	28.04	0.787	100	1.11E+06	0.897
<b>13C12-1,2,3,7,8-PeCDD</b>	32.1	1.57	100	8.54E+05	0.688
<b>13C12-1,2,3,6,7,8-HxCDD</b>	34.2	1.247	100	8.54E+05	1.088
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>	35.82	1.072	100	6.08E+05	0.775
<b>13C12-OCDD</b>	37.32	0.884	200	1.01E+06	0.641
<b>13C12-2,3,7,8-TCDF</b>	27.11	0.747	100	1.76E+06	1.421
<b>13C12-1,2,3,7,8-PeCDF</b>	31.17	1.601	100	1.24E+06	1
<b>13C12-1,2,3,6,7,8-HxCDF</b>	33.73	0.532	100	9.90E+05	1.263
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>	35.26	0.443	100	6.94E+05	0.885
<b>Cleanup Standard</b>					
<b>13C12-1,2,3,7,8,9-HxCDF</b>	34.48	0.511	100	8.07E+05	1.029
<b>Injection Standards</b>					
<b>13C12-1234-TCDD IS</b>	27.35	0.791	100	1240529.3	12405.293
<b>13C12-123789-HxCDD IS</b>	34.33	1.25	100.00	7.84E+05	7841.906

# ALS Life Sciences

## Calibration Report

ALS Sample ID **H7-19-CCV-CS2-0017**  
 Analysis Method EPA TO9A  
 Analysis Type Calibration

Filename 7-190724A02    Inst # HRMS-7    Column DB5MSUSR826231H    Run Date 24-Jul-2019 16:37

Approved: *T.Patterson*  
 --e-signature--  
 19-Aug-2019

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
<b>2,3,7,8-TCDD</b>	28.08	0.78	2.00	1.71E+04	0.999
<b>1,2,3,7,8-PeCDD</b>	32.12	1.67	10.00	6.17E+04	0.895
<b>1,2,3,4,7,8-HxCDD</b>	34.15	1.25	10.00	5.75E+04	0.841
<b>1,2,3,6,7,8-HxCDD</b>	34.21	1.27	10.00	6.25E+04	0.914
<b>1,2,3,7,8,9-HxCDD</b>	34.33	1.25	10.00	6.02E+04	0.879
<b>1,2,3,4,6,7,8-HpCDD</b>	35.82	1.03	10.00	4.88E+04	0.884
<b>OCDD</b>	37.32	0.85	20.00	8.97E+04	0.937
<b>2,3,7,8-TCDF</b>	27.16	0.77	2.00	2.13E+04	0.788
<b>1,2,3,7,8-PeCDF</b>	31.19	1.56	10.00	8.29E+04	0.834
<b>2,3,4,7,8-PeCDF</b>	31.89	1.55	10.00	9.34E+04	0.940
<b>1,2,3,4,7,8-HxCDF</b>	33.66	1.21	10.00	7.66E+04	0.896
<b>1,2,3,6,7,8-HxCDF</b>	33.73	1.16	10.00	8.24E+04	0.964
<b>2,3,4,6,7,8-HxCDF</b>	34.06	1.21	10.00	7.69E+04	0.899
<b>1,2,3,7,8,9-HxCDF</b>	34.48	1.23	10.00	6.41E+04	0.749
<b>1,2,3,4,6,7,8-HpCDF</b>	35.27	1.94	10.00	5.29E+04	0.849
<b>1,2,3,4,7,8,9-HpCDF</b>	36.07	1.93	10.00	4.33E+04	0.695
<b>OCDF</b>	37.4	0.88	20.00	1.04E+05	1.081
<b>Field Spike Standards</b>					
<b>37Cl4-2,3,7,8-TCDD</b>	28.08	0.00	2.00	1.85E+04	1.082
<b>13C12-1,2,3,4,7,8-HxCDD</b>	34.15	1.26	100.00	6.23E+05	0.911
<b>13C12-2,3,4,7,8-PeCDF</b>	31.88	1.57	100.00	1.02E+06	1.025
<b>13C12-1,2,3,4,7,8-HxCDF</b>	33.66	0.52	100.00	7.26E+05	0.849
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>	36.07	0.43	100.00	4.99E+05	0.802
<b>Extraction Standards</b>					
<b>13C12-2,3,7,8-TCDD</b>	28.05	0.793	100	8.55E+05	0.899
<b>13C12-1,2,3,7,8-PeCDD</b>	32.1	1.582	100	6.90E+05	0.726
<b>13C12-1,2,3,6,7,8-HxCDD</b>	34.2	1.261	100	6.84E+05	1.109
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>	35.81	1.012	100	5.51E+05	0.893
<b>13C12-OCDD</b>	37.31	0.884	200	9.58E+05	0.776
<b>13C12-2,3,7,8-TCDF</b>	27.13	0.773	100	1.35E+06	1.421
<b>13C12-1,2,3,7,8-PeCDF</b>	31.18	1.481	100	9.94E+05	1.046
<b>13C12-1,2,3,6,7,8-HxCDF</b>	33.73	0.538	100	8.55E+05	1.385
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>	35.26	0.443	100	6.22E+05	1.008
<b>Cleanup Standard</b>					
<b>13C12-1,2,3,7,8,9-HxCDF</b>	34.47	0.512	100	6.92E+05	1.121
<b>Injection Standards</b>					
<b>13C12-1234-TCDD IS</b>	27.38	0.787	100	950730.5	9507.305
<b>13C12-123789-HxCDD IS</b>	34.33	1.27	100.00	6.17E+05	6172.903

# ALS Life Sciences

## Calibration Report

ALS Sample ID **H7-19-CCV-CS3-0017**  
 Analysis Method EPA TO9A  
 Analysis Type Calibration

Filename 7-190724A01    Inst # HRMS-7    Column DB5MSUSR826231H    Run Date 24-Jul-2019 15:56

Approved: *T.Patterson*  
 --e-signature--  
 19-Aug-2019

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
<b>2,3,7,8-TCDD</b>	28.06	0.76	10.00	1.43E+05	1.261
<b>1,2,3,7,8-PeCDD</b>	32.12	1.66	50.00	4.18E+05	0.923
<b>1,2,3,4,7,8-HxCDD</b>	34.15	1.22	50.00	3.88E+05	0.871
<b>1,2,3,6,7,8-HxCDD</b>	34.21	1.23	50.00	4.19E+05	0.940
<b>1,2,3,7,8,9-HxCDD</b>	34.34	1.19	50.00	4.05E+05	0.909
<b>1,2,3,4,6,7,8-HpCDD</b>	35.82	1.03	50.00	3.32E+05	0.946
<b>OCDD</b>	37.32	0.88	100.00	5.83E+05	0.938
<b>2,3,7,8-TCDF</b>	27.14	0.75	10.00	1.67E+05	0.913
<b>1,2,3,7,8-PeCDF</b>	31.19	1.52	50.00	5.65E+05	0.864
<b>2,3,4,7,8-PeCDF</b>	31.9	1.52	50.00	6.22E+05	0.950
<b>1,2,3,4,7,8-HxCDF</b>	33.67	1.19	50.00	5.03E+05	0.911
<b>1,2,3,6,7,8-HxCDF</b>	33.74	1.16	50.00	5.48E+05	0.993
<b>2,3,4,6,7,8-HxCDF</b>	34.06	1.17	50.00	5.08E+05	0.920
<b>1,2,3,7,8,9-HxCDF</b>	34.48	1.22	50.00	4.52E+05	0.818
<b>1,2,3,4,6,7,8-HpCDF</b>	35.27	1.84	50.00	3.45E+05	0.888
<b>1,2,3,4,7,8,9-HpCDF</b>	36.08	1.81	50.00	2.97E+05	0.764
<b>OCDF</b>	37.41	0.89	100.00	6.72E+05	1.082
<b>Field Spike Standards</b>					
<b>37Cl4-2,3,7,8-TCDD</b>	28.06	0.00	10.00	1.37E+05	1.203
<b>13C12-1,2,3,4,7,8-HxCDD</b>	34.15	1.36	100.00	8.05E+05	0.903
<b>13C12-2,3,4,7,8-PeCDF</b>	31.89	1.55	100.00	1.36E+06	1.043
<b>13C12-1,2,3,4,7,8-HxCDF</b>	33.66	0.51	100.00	9.46E+05	0.856
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>	36.07	0.44	100.00	6.32E+05	0.813
<b>Extraction Standards</b>					
<b>13C12-2,3,7,8-TCDD</b>	28.05	0.804	100	1.14E+06	0.909
<b>13C12-1,2,3,7,8-PeCDD</b>	32.1	1.595	100	9.06E+05	0.725
<b>13C12-1,2,3,6,7,8-HxCDD</b>	34.21	1.191	100	8.91E+05	1.099
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>	35.81	1.046	100	7.02E+05	0.865
<b>13C12-OCDD</b>	37.31	0.905	200	1.24E+06	0.766
<b>13C12-2,3,7,8-TCDF</b>	27.13	0.754	100	1.83E+06	1.462
<b>13C12-1,2,3,7,8-PeCDF</b>	31.18	1.586	100	1.31E+06	1.047
<b>13C12-1,2,3,6,7,8-HxCDF</b>	33.73	0.519	100	1.10E+06	1.362
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>	35.26	0.446	100	7.77E+05	0.957
<b>Cleanup Standard</b>					
<b>13C12-1,2,3,7,8,9-HxCDF</b>	34.47	0.524	100	8.99E+05	1.108
<b>Injection Standards</b>					
<b>13C12-1234-TCDD IS</b>	27.36	0.809	100	1250070.6	12500.706
<b>13C12-123789-HxCDD IS</b>	34.33	1.23	100.00	8.11E+05	8111.065

# ALS Life Sciences

## Calibration Report

ALS Sample ID **H7-19-CCV-CS4-0017**  
 Analysis Method EPA TO9A  
 Analysis Type Calibration

Filename 7-190724A07      Inst # HRMS-7      Column DB5MSUSR826231H      Run Date 24-Jul-2019 20:08

Approved: *T.Patterson*  
 --e-signature--  
 19-Aug-2019

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
<b>2,3,7,8-TCDD</b>	28.04	0.77	40.00	5.30E+05	1.155
<b>1,2,3,7,8-PeCDD</b>	32.1	1.64	200.00	1.79E+06	0.948
<b>1,2,3,4,7,8-HxCDD</b>	34.15	1.31	200.00	1.73E+06	0.905
<b>1,2,3,6,7,8-HxCDD</b>	34.2	1.17	200.00	1.92E+06	1.005
<b>1,2,3,7,8,9-HxCDD</b>	34.32	1.23	200.00	1.88E+06	0.981
<b>1,2,3,4,6,7,8-HpCDD</b>	35.81	1.03	200.00	1.44E+06	1.000
<b>OCDD</b>	37.31	0.88	400.00	2.60E+06	1.010
<b>2,3,7,8-TCDF</b>	27.11	0.76	40.00	6.49E+05	0.904
<b>1,2,3,7,8-PeCDF</b>	31.17	1.55	200.00	2.39E+06	0.908
<b>2,3,4,7,8-PeCDF</b>	31.88	1.52	200.00	2.73E+06	1.035
<b>1,2,3,4,7,8-HxCDF</b>	33.66	1.18	200.00	2.18E+06	0.965
<b>1,2,3,6,7,8-HxCDF</b>	33.73	1.18	200.00	2.36E+06	1.041
<b>2,3,4,6,7,8-HxCDF</b>	34.05	1.17	200.00	2.20E+06	0.972
<b>1,2,3,7,8,9-HxCDF</b>	34.47	1.17	200.00	1.98E+06	0.874
<b>1,2,3,4,6,7,8-HpCDF</b>	35.26	1.86	200.00	1.50E+06	0.936
<b>1,2,3,4,7,8,9-HpCDF</b>	36.07	1.86	200.00	1.26E+06	0.786
<b>OCDF</b>	37.39	0.89	400.00	3.05E+06	1.185
<b>Field Spike Standards</b>					
<b>37Cl4-2,3,7,8-TCDD</b>	28.04	0.00	40.00	5.26E+05	1.145
<b>13C12-1,2,3,4,7,8-HxCDD</b>	34.14	1.33	100.00	8.71E+05	0.911
<b>13C12-2,3,4,7,8-PeCDF</b>	31.87	1.56	100.00	1.41E+06	1.067
<b>13C12-1,2,3,4,7,8-HxCDF</b>	33.65	0.52	100.00	9.93E+05	0.878
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>	36.05	0.45	100.00	6.68E+05	0.832
<b>Extraction Standards</b>					
<b>13C12-2,3,7,8-TCDD</b>	28.01	0.78	100	1.15E+06	0.908
<b>13C12-1,2,3,7,8-PeCDD</b>	32.09	1.564	100	9.45E+05	0.748
<b>13C12-1,2,3,6,7,8-HxCDD</b>	34.19	1.144	100	9.56E+05	1.083
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>	35.8	1.033	100	7.20E+05	0.815
<b>13C12-OCDD</b>	37.3	0.887	200	1.29E+06	0.729
<b>13C12-2,3,7,8-TCDF</b>	27.1	0.786	100	1.80E+06	1.421
<b>13C12-1,2,3,7,8-PeCDF</b>	31.16	1.571	100	1.32E+06	1.042
<b>13C12-1,2,3,6,7,8-HxCDF</b>	33.71	0.517	100	1.13E+06	1.28
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>	35.24	0.447	100	8.03E+05	0.909
<b>Cleanup Standard</b>					
<b>13C12-1,2,3,7,8,9-HxCDF</b>	34.46	0.525	100	9.58E+05	1.085
<b>Injection Standards</b>					
<b>13C12-1234-TCDD IS</b>	27.34	0.798	100	1264239.9	12642.399
<b>13C12-123789-HxCDD IS</b>	34.32	1.15	100.00	8.83E+05	8834.065

# ALS Life Sciences

## Calibration Report

ALS Sample ID **H7-19-CCV-CS5-0017**  
 Analysis Method EPA TO9A  
 Analysis Type Calibration

Filename 7-190724A06	Inst # HRMS-7	Column DB5MSUSR826231H	Run Date 24-Jul-2019 19:26	Approved: <i>T.Patterson</i> --e-signature-- 19-Aug-2019
-------------------------	------------------	---------------------------	-------------------------------	--

Target Analytes	Ret. Time	Ion Ratio	Concentration ng/mL	Response	RRF
<b>2,3,7,8-TCDD</b>					
28.04 0.79 200.00 2.33E+06 1.148					
<b>1,2,3,7,8-PeCDD</b>					
32.1 1.62 1000.00 8.37E+06 0.914					
<b>1,2,3,4,7,8-HxCDD</b>					
34.15 1.23 1000.00 8.38E+06 0.896					
<b>1,2,3,6,7,8-HxCDD</b>					
34.2 1.23 1000.00 8.83E+06 0.944					
<b>1,2,3,7,8,9-HxCDD</b>					
34.32 1.23 1000.00 8.48E+06 0.906					
<b>1,2,3,4,6,7,8-HpCDD</b>					
35.81 1.03 1000.00 7.13E+06 0.965					
<b>OCDD</b>					
37.31 0.88 2000.00 1.41E+07 0.962					
<b>2,3,7,8-TCDF</b>					
27.11 0.76 200.00 2.77E+06 0.870					
<b>1,2,3,7,8-PeCDF</b>					
31.17 1.54 1000.00 1.13E+07 0.881					
<b>2,3,4,7,8-PeCDF</b>					
31.88 1.53 1000.00 1.27E+07 0.988					
<b>1,2,3,4,7,8-HxCDF</b>					
33.66 1.17 1000.00 1.04E+07 0.937					
<b>1,2,3,6,7,8-HxCDF</b>					
33.73 1.18 1000.00 1.10E+07 0.984					
<b>2,3,4,6,7,8-HxCDF</b>					
34.05 1.18 1000.00 1.05E+07 0.943					
<b>1,2,3,7,8,9-HxCDF</b>					
34.47 1.18 1000.00 9.51E+06 0.853					
<b>1,2,3,4,6,7,8-HpCDF</b>					
35.26 1.84 1000.00 7.27E+06 0.910					
<b>1,2,3,4,7,8,9-HpCDF</b>					
36.05 1.86 1000.00 6.30E+06 0.788					
<b>OCDF</b>					
37.39 0.89 2000.00 1.67E+07 1.137					
<b>Field Spike Standards</b>					
<b>37Cl4-2,3,7,8-TCDD</b>					
28.04 0.00 200.00 2.56E+06 1.262					
<b>13C12-1,2,3,4,7,8-HxCDD</b>					
34.14 1.25 100.00 8.78E+05 0.938					
<b>13C12-2,3,4,7,8-PeCDF</b>					
31.87 1.56 100.00 1.35E+06 1.049					
<b>13C12-1,2,3,4,7,8-HxCDF</b>					
33.65 0.51 100.00 9.65E+05 0.866					
<b>13C12-1,2,3,4,7,8,9-HpCDF</b>					
36.05 0.44 100.00 6.85E+05 0.856					
<b>Extraction Standards</b>					
<b>13C12-2,3,7,8-TCDD</b>					
28.02 0.785 100 1.02E+06 0.934					
<b>13C12-1,2,3,7,8-PeCDD</b>					
32.09 1.602 100 9.15E+05 0.842					
<b>13C12-1,2,3,6,7,8-HxCDD</b>					
34.19 1.255 100 9.36E+05 1.094					
<b>13C12-1,2,3,4,6,7,8-HpCDD</b>					
35.8 1.04 100 7.39E+05 0.863					
<b>13C12-OCDD</b>					
37.3 0.897 200 1.47E+06 0.857					
<b>13C12-2,3,7,8-TCDF</b>					
27.1 0.758 100 1.59E+06 1.467					
<b>13C12-1,2,3,7,8-PeCDF</b>					
31.16 1.548 100 1.28E+06 1.181					
<b>13C12-1,2,3,6,7,8-HxCDF</b>					
33.72 0.501 100 1.11E+06 1.303					
<b>13C12-1,2,3,4,6,7,8-HpCDF</b>					
35.24 0.445 100 8.00E+05 0.934					
<b>Cleanup Standard</b>					
<b>13C12-1,2,3,7,8,9-HxCDF</b>					
34.46 0.514 100 9.49E+05 1.109					
<b>Injection Standards</b>					
<b>13C12-1234-TCDD IS</b>					
27.35 0.788 100 1086981.5 10869.815					
<b>13C12-123789-HxCDD IS</b>					
34.32 1.24 100.00 8.56E+05 8558.793					

# ALS Life Sciences

## Second Source Calibration Verification Report

<b>Sample Name</b>	CVS	Sampling Date	n/a	
ALS Sample ID	1613B-RS1-011	Extraction Date	n/a	Approved: <i>S. Kennedy</i> --e-signature-- 15-Aug-2019
Analysis Method	EPA 1613B	Sample Size	1 n/a	
Analysis Type	CCV	Percent Moisture	n/a	
Sample Matrix	QC	Split Ratio	1	

<b>Run Information</b>	<b>Run 1</b>
Filename	7-190724A08
Run Date	24-Jul-19 20:50
Final Volume	20 uL
Dilution Factor	1
Analysis Units	%
Instrument - Column	HRMS-7 DB5MSUSR826231H

Target Analytes	pg/uL	Ret. Time	% Rec	Limits	Flags
2,3,7,8-TCDD	10	28.09	104.3	78-129	
1,2,3,7,8-PeCDD	50	32.13	109.9	78-130	
1,2,3,4,7,8-HxCDD	50	34.16	100.2	78-128	
1,2,3,6,7,8-HxCDD	50	34.21	97.9	78-128	
1,2,3,7,8,9-HxCDD	50	34.34	105.7	82-122	
1,2,3,4,6,7,8-HpCDD	50	35.82	99.6	86-116	
OCDD	100	37.32	98	79-126	
2,3,7,8-TCDF	10	27.17	97.6	84-120	
1,2,3,7,8-PeCDF	50	31.19	106.6	82-120	
2,3,4,7,8-PeCDF	50	31.90	94.9	82-122	
1,2,3,4,7,8-HxCDF	50	33.67	99.5	90-112	
1,2,3,6,7,8-HxCDF	50	33.74	105.7	88-114	
2,3,4,6,7,8-HxCDF	50	34.06	99.4	88-114	
1,2,3,7,8,9-HxCDF	50	34.48	108.2	90-112	
1,2,3,4,6,7,8-HpCDF	50	35.27	98.8	90-110	
1,2,3,4,7,8,9-HpCDF	50	36.08	103.1	86-116	
OCDF	100	37.41	103.9	63-159	
<b>Extraction Standards</b>	<b>pg/uL</b>		<b>% Rec</b>	<b>Limits</b>	
13C12-2,3,7,8-TCDD	100	28.06	99.2	82-121	
13C12-1,2,3,7,8-PeCDD	100	32.12	104.3	62-160	
13C12-1,2,3,4,7,8-HxCDD	100	34.15	93.2	85-117	
13C12-1,2,3,6,7,8-HxCDD	100	34.21	96.9	85-118	
13C12-1,2,3,4,6,7,8-HpCDD	100	35.82	96.6	72-138	
13C12-OCDD	200	37.32	90.5	48-207	
13C12-2,3,7,8-TCDF	100	27.14	99.3	71-140	
13C12-1,2,3,7,8-PeCDF	100	31.18	101.8	76-130	
13C12-2,3,4,7,8-PeCDF	100	31.89	101	77-130	
13C12-1,2,3,4,7,8-HxCDF	100	33.66	94.7	76-131	
13C12-1,2,3,6,7,8-HxCDF	100	33.73	93.2	70-143	
13C12-2,3,4,6,7,8-HxCDF	100	34.06	95.3	73-137	
13C12-1,2,3,7,8,9-HxCDF	100	34.48	92.3	74-135	
13C12-1,2,3,4,6,7,8-HpCDF	100	35.27	95.5	78-129	
13C12-1,2,3,4,7,8,9-HpCDF	100	36.08	95.4	77-129	
<b>Cleanup Standard</b>	<b>pg/uL</b>				
37C14-2,3,7,8-TCDD (Cleanup)	10	28.09	100.4	31-191	

# ALS Life Sciences

## Continuing Calibration Report

<b>Sample Name</b>	CCV	Sampling Date	n/a	
ALS Sample ID	H7-19-CCV-0534	Extraction Date	n/a	
Analysis Method	EPA TO9A	Sample Size	1	n/a
Analysis Type	CCV	Percent Moisture	n/a	
Sample Matrix	QC	Split Ratio	1	

Approved:  
T.Patterson  
--e-signature--  
19-Aug-2019

<b>Run Information</b>	<b>Run 1</b>
Filename	7-190816A07
Run Date	16-Aug-19 02:13
Final Volume	10 uL
Dilution Factor	1
Analysis Units	%
Instrument - Column	HRMS-7 DB5MSUSR826231H

Target Analytes	pg/uL	Ret. Time	% Rec	Limits	Flags
2,3,7,8-TCDD	10	28.12	122	75-125	
1,2,3,7,8-PeCDD	50	32.16	114	75-125	
1,2,3,4,7,8-HxCDD	50	34.21	96	75-125	
1,2,3,6,7,8-HxCDD	50	34.26	111	75-125	
1,2,3,7,8,9-HxCDD	50	34.39	99	75-125	
1,2,3,4,6,7,8-HpCDD	50	35.89	110	75-125	
OCDD	100	37.39	110	75-125	
2,3,7,8-TCDF	10	27.20	125	75-125	
1,2,3,7,8-PeCDF	50	31.24	109	75-125	
2,3,4,7,8-PeCDF	50	31.95	125	75-125	
1,2,3,4,7,8-HxCDF	50	33.72	98	75-125	
1,2,3,6,7,8-HxCDF	50	33.79	114	75-125	
2,3,4,6,7,8-HxCDF	50	34.12	108	75-125	
1,2,3,7,8,9-HxCDF	50	34.54	104	75-125	
1,2,3,4,6,7,8-HpCDF	50	35.32	103	75-125	
1,2,3,4,7,8,9-HpCDF	50	36.13	105	75-125	
OCDF	100	37.48	121	70-130	
<b>Field Spike Standards</b>	<b>pg/uL</b>		<b>% Rec</b>	<b>Limits</b>	
37C14-2,3,7,8-TCDD	10	28.12	107	75-125	
13C12-1,2,3,4,7,8-HxCDD	100	34.21	86	75-125	
13C12-2,3,4,7,8-PeCDF	100	31.94	110	75-125	
13C12-1,2,3,4,7,8-HxCDF	100	33.71	89	75-125	
13C12-1,2,3,4,7,8,9-HpCDF	100	36.13	95	75-125	
<b>Extraction Standards</b>					
13C12-2,3,7,8-TCDD	100	28.11	105	75-125	
13C12-1,2,3,7,8-PeCDD	100	32.15	105	70-130	
13C12-1,2,3,6,7,8-HxCDD	100	34.26	113	75-125	
13C12-1,2,3,4,6,7,8-HpCDD	100	35.88	122	70-130	
13C12-OCDD	200	37.38	111	70-130	
13C12-2,3,7,8-TCDF	100	27.19	129	70-130	
13C12-1,2,3,7,8-PeCDF	100	31.23	128	70-130	
13C12-1,2,3,6,7,8-HxCDF	100	33.78	119	70-130	
13C12-1,2,3,4,6,7,8-HpCDF	100	35.32	145	70-130	
<b>Cleanup Standard</b>	<b>pg/uL</b>				
13C12-1,2,3,7,8,9-HxCDF	100	34.53	110	40-130	



# ALS Life Sciences

## Continuing Calibration Report

<b>Sample Name</b>	CCV	Sampling Date	n/a		
ALS Sample ID	H7-19-CCV-0535	Extraction Date	n/a		
Analysis Method	EPA T09A	Sample Size	1	n/a	
Analysis Type	CCV	Percent Moisture	n/a		
Sample Matrix	QC	Split Ratio	1		

Approved:  
T.Patterson  
--e-signature--  
19-Aug-2019

<b>Run Information</b>	<b>Run 1</b>
Filename	7-190816A25
Run Date	16-Aug-19 14:59
Final Volume	10 uL
Dilution Factor	1
Analysis Units	%
Instrument - Column	HRMS-7 DB5MSUSR826231H

Target Analytes	pg/uL	Ret. Time	% Rec	Limits	Flags
2,3,7,8-TCDD	10	28.15	121	75-125	
1,2,3,7,8-PeCDD	50	32.17	110	75-125	
1,2,3,4,7,8-HxCDD	50	34.21	91	75-125	
1,2,3,6,7,8-HxCDD	50	34.26	104	75-125	
1,2,3,7,8,9-HxCDD	50	34.39	95	75-125	
1,2,3,4,6,7,8-HpCDD	50	35.89	113	75-125	
OCDD	100	37.39	108	75-125	
2,3,7,8-TCDF	10	27.23	125	75-125	
1,2,3,7,8-PeCDF	50	31.24	112	75-125	
2,3,4,7,8-PeCDF	50	31.95	114	75-125	
1,2,3,4,7,8-HxCDF	50	33.72	88	75-125	
1,2,3,6,7,8-HxCDF	50	33.79	109	75-125	
2,3,4,6,7,8-HxCDF	50	34.12	111	75-125	
1,2,3,7,8,9-HxCDF	50	34.54	110	75-125	
1,2,3,4,6,7,8-HpCDF	50	35.32	104	75-125	
1,2,3,4,7,8,9-HpCDF	50	36.13	104	75-125	
OCDF	100	37.48	117	70-130	
<b>Field Spike Standards</b>					
37C14-2,3,7,8-TCDD	10	28.15	111	75-125	
13C12-1,2,3,4,7,8-HxCDD	100	34.21	84	75-125	
13C12-2,3,4,7,8-PeCDF	100	31.94	100	75-125	
13C12-1,2,3,4,7,8-HxCDF	100	33.71	84	75-125	
13C12-1,2,3,4,7,8,9-HpCDF	100	36.13	95	75-125	
<b>Extraction Standards</b>					
13C12-2,3,7,8-TCDD	100	28.12	96	75-125	
13C12-1,2,3,7,8-PeCDD	100	32.15	96	70-130	
13C12-1,2,3,6,7,8-HxCDD	100	34.25	102	75-125	
13C12-1,2,3,4,6,7,8-HpCDD	100	35.88	104	70-130	
13C12-OCDD	200	37.38	112	70-130	
13C12-2,3,7,8-TCDF	100	27.22	101	70-130	
13C12-1,2,3,7,8-PeCDF	100	31.23	108	70-130	
13C12-1,2,3,6,7,8-HxCDF	100	33.78	109	70-130	
13C12-1,2,3,4,6,7,8-HpCDF	100	35.32	125	70-130	
<b>Cleanup Standard</b>					
13C12-1,2,3,7,8,9-HxCDF	100	34.53	108	40-130	

## **SVOC DATA PACKAGE**

### **SECTION 5: QC SAMPLE DATA**

Including:

- Laboratory Method Blank Analysis Reports
- Laboratory Control Sample Analysis Reports
- Matrix Spike Analysis Reports
- Other QC Sample Analysis Reports (where applicable)

# ALS Life Sciences

## Laboratory Method Blank Analysis Report

**Sample Name**                      **Method Blank**  
 ALS Sample ID                    WG3116369-1  
 Analysis Method                EPA TO9A  
 Analysis Type                    Blank  
 Sample Matrix                  MEDIA

Sampling Date                    n/a  
 Extraction Date                12-Aug-19  
 Sample Size                    1                    Puf  
 Percent Moisture               n/a  
 Split Ratio                      2

Approved:  
*T.Patterson*  
 --e-signature--  
 19-Aug-2019

**Run Information**                      **Run 1**  
 Filename                            7-190816A11  
 Run Date                            16-Aug-19 05:10  
 Final Volume                      10                    uL  
 Dilution Factor                  1  
 Analysis Units                    pg  
 Instrument - Column              HRMS-7 DB5MSUSR826231H

Target Analytes	TEF (WHO 2005)	Ret. Time	Conc. pg	EDL pg	Flags	EMPC pg	LQL
2,3,7,8-TCDD	1	NotFnd	<2.3	2.3	U		10
1,2,3,7,8-PeCDD	1	NotFnd	<1.3	1.3	U		50
1,2,3,4,7,8-HxCDD	0.1	NotFnd	<1.4	1.4	U		50
1,2,3,6,7,8-HxCDD	0.1	NotFnd	<1.3	1.3	U		50
1,2,3,7,8,9-HxCDD	0.1	NotFnd	<1.3	1.3	U		50
1,2,3,4,6,7,8-HpCDD	0.01	NotFnd	<2.3	2.3	U		50
OCDD	0.0003	37.39	<6.2	2.5	M,I,R	6.2	100
2,3,7,8-TCDF	0.1	NotFnd	<1.5	1.5	U		10
1,2,3,7,8-PeCDF	0.03	NotFnd	<1.1	1.1	U		50
2,3,4,7,8-PeCDF	0.3	NotFnd	<0.97	0.97	U		50
1,2,3,4,7,8-HxCDF	0.1	NotFnd	<0.97	0.97	U		50
1,2,3,6,7,8-HxCDF	0.1	NotFnd	<0.90	0.90	U		50
2,3,4,6,7,8-HxCDF	0.1	NotFnd	<0.97	0.97	U		50
1,2,3,7,8,9-HxCDF	0.1	34.54	1.11	1.1	M,J		50
1,2,3,4,6,7,8-HpCDF	0.01	NotFnd	<1.2	1.2	U		50
1,2,3,4,7,8,9-HpCDF	0.01	NotFnd	<1.4	1.4	U		50
OCDF	0.0003	37.49	3.42	2.3	M,J		100

**Field Spike Standards**                      **% Rec**

37Cl4-2,3,7,8-TCDD	NS
13C12-1,2,3,4,7,8-HxCDD	NS
13C12-2,3,4,7,8-PeCDF	NS
13C12-1,2,3,4,7,8-HxCDF	NS
13C12-1,2,3,4,7,8,9-HpCDF	NS

**Extraction Standards**

13C12-2,3,7,8-TCDD	4000	28.11	48	40-130
13C12-1,2,3,7,8-PeCDD	4000	32.15	60	40-130
13C12-1,2,3,6,7,8-HxCDD	4000	34.26	56	40-130
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.88	59	25-130
13C12-OCDD	8000	37.38	48	25-130
13C12-2,3,7,8-TCDF	4000	27.19	67	40-130
13C12-1,2,3,7,8-PeCDF	4000	31.23	68	40-130
13C12-1,2,3,6,7,8-HxCDF	4000	33.79	69	40-130
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.32	71	25-130

**Cleanup Standard**                      **pg**

13C12-1,2,3,7,8,9-HxCDF	4000	34.53	66	40-130
-------------------------	------	-------	----	--------

**Homologue Group Totals**                      **# peaks**                      **Conc. pg**                      **EDL pg**

Total-TCDD	0	<2.3	2.3	U	10
Total-PeCDD	0	<1.3	1.3	U	50
Total-HxCDD	0	<1.4	1.4	U	50
Total-HpCDD	0	<2.3	2.3	U	50
Total-TCDF	0	<1.5	1.5	U	10
Total-PeCDF	0	<1.1	1.1	U	50
Total-HxCDF	1	1.11	1.1		50
Total-HpCDF	0	<1.4	1.4	U	50

**Toxic Equivalency - (WHO 2005)**                      **pg**

<b>Lower Bound PCDD/F TEQ (WHO 2005)</b>	0.112
<b>Mid Point PCDD/F TEQ (WHO 2005)</b>	2.52
<b>Upper Bound PCDD/F TEQ (WHO 2005)</b>	4.92

EDL                      Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.  
 TEF                      Indicates the Toxic Equivalency Factor                      TEQ                      Indicates the Toxic Equivalency  
 M                      Indicates that a peak has been manually integrated.  
 U                      Indicates that this compound was not detected above the EDL.  
  
 J                      Indicates that a target analyte was detected below the calibrated range.  
 R                      Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.  
  
 LQL                      Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.  
 EMPC                      Estimated Maximum Possible Concentration - elevated detection limit due to interference or positive id criterion failure  
 NS                      Indicates that this standard was not spiked to sample

# ALS Life Sciences

## Laboratory Method Blank Analysis Report

<b>Sample Name</b>	Method Blank	Sampling Date	n/a		
ALS Sample ID	WG3116369-4	Extraction Date	12-Aug-19		
Analysis Method	EPA TO9A	Sample Size	1	Puf	
Analysis Type	Blank	Percent Moisture	n/a		Approved: T. Patterson --e-signature-- 19-Aug-2019
Sample Matrix	REAGENT	Split Ratio	2		

<b>Run Information</b>	<b>Run 1</b>
Filename	7-190816A12
Run Date	16-Aug-19 05:52
Final Volume	10 uL
Dilution Factor	1
Analysis Units	pg
Instrument - Column	HRMS-7 DB5MSUSR826231H

Target Analytes	TEF (WHO 2005)	Ret. Time	Conc. pg	EDL pg	Flags	EMPC pg	LQL
2,3,7,8-TCDD	1	NotFnd	<2.0	2.0	U		10
1,2,3,7,8-PeCDD	1	NotFnd	<1.1	1.1	U		50
1,2,3,4,7,8-HxCDD	0.1	NotFnd	<1.5	1.5	U		50
1,2,3,6,7,8-HxCDD	0.1	NotFnd	<1.3	1.3	U		50
1,2,3,7,8,9-HxCDD	0.1	NotFnd	<1.4	1.4	U		50
1,2,3,4,6,7,8-HpCDD	0.01	35.89	<2.1	1.1	M,J,R	2.1	50
OCDD	0.0003	37.40	8.95	1.1	J		100
2,3,7,8-TCDF	0.1	NotFnd	<1.4	1.4	U		10
1,2,3,7,8-PeCDF	0.03	NotFnd	<0.82	0.82	U		50
2,3,4,7,8-PeCDF	0.3	NotFnd	<0.74	0.74	U		50
1,2,3,4,7,8-HxCDF	0.1	NotFnd	<0.79	0.79	U		50
1,2,3,6,7,8-HxCDF	0.1	NotFnd	<0.73	0.73	U		50
2,3,4,6,7,8-HxCDF	0.1	NotFnd	<0.78	0.78	U		50
1,2,3,7,8,9-HxCDF	0.1	34.55	<0.93	0.89	M,J,R	0.93	50
1,2,3,4,6,7,8-HpCDF	0.01	NotFnd	<0.76	0.76	U		50
1,2,3,4,7,8,9-HpCDF	0.01	NotFnd	<0.90	0.90	U		50
OCDF	0.0003	37.49	2.42	1.1	M,J,B		100

<b>Field Spike Standards</b>	<b>% Rec</b>
37Cl4-2,3,7,8-TCDD	NS
13C12-1,2,3,4,7,8-HxCDD	NS
13C12-2,3,4,7,8-PeCDF	NS
13C12-1,2,3,4,7,8-HxCDF	NS
13C12-1,2,3,4,7,8,9-HpCDF	NS

<b>Extraction Standards</b>							
13C12-2,3,7,8-TCDD	4000	28.12	52	40-130			
13C12-1,2,3,7,8-PeCDD	4000	32.16	68	40-130			
13C12-1,2,3,6,7,8-HxCDD	4000	34.26	59	40-130			
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.89	67	25-130			
13C12-OCDD	8000	37.39	54	25-130			
13C12-2,3,7,8-TCDF	4000	27.20	67	40-130			
13C12-1,2,3,7,8-PeCDF	4000	31.23	73	40-130			
13C12-1,2,3,6,7,8-HxCDF	4000	33.79	62	40-130			
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.33	74	25-130			

<b>Cleanup Standard</b>	<b>pg</b>						
13C12-1,2,3,7,8,9-HxCDF	4000	34.54	70	40-130			

<b>Homologue Group Totals</b>	<b># peaks</b>	<b>Conc.</b>	<b>EDL</b>				
Total-TCDD	0	<2.0	2.0	U			10
Total-PeCDD	0	<1.1	1.1	U			50
Total-HxCDD	0	<1.5	1.5	U			50
Total-HpCDD	0	<1.1	1.1	U			50
Total-TCDF	0	<1.4	1.4	U			10
Total-PeCDF	0	<0.82	0.82	U			50
Total-HxCDF	0	<0.89	0.89	U			50
Total-HpCDF	0	<0.90	0.90	U			50

<b>Toxic Equivalency - (WHO 2005)</b>	<b>pg</b>
<b>Lower Bound PCDD/F TEQ (WHO 2005)</b>	0.00341
<b>Mid Point PCDD/F TEQ (WHO 2005)</b>	2.19
<b>Upper Bound PCDD/F TEQ (WHO 2005)</b>	4.27

EDL	Indicates the Estimated Detection Limit, based on the measured background noise for this target in this sample.
TEF	Indicates the Toxic Equivalency Factor      TEQ      Indicates the Toxic Equivalency
M	Indicates that a peak has been manually integrated.
U	Indicates that this compound was not detected above the EDL.
J	Indicates that a target analyte was detected below the calibrated range.
R	Indicates that the ion abundance ratio for this compound did not meet the acceptance criterion.
B	Indicates that this target was detected in the blank at greater than 10% of the sample concentration.
LQL	Lower Quantification Limit, based on the lowest calibration level corrected for sample size, splits and dilutions.
EMPC	Estimated Maximum Possible Concentration - elevated detection limit due to interference or positive id criterion failure
NS	Indicates that this standard was not spiked to sample

# ALS Life Sciences

## Laboratory Control Sample Analysis Report

<b>Sample Name</b>	<b>Laboratory Control Sample</b>	Sampling Date	n/a
ALS Sample ID	WG3116369-2	Extraction Date	12-Aug-19
Analysis Method	EPA TO9A	Sample Size	1 n/a
Analysis Type	LCS	Percent Moisture	n/a
Sample Matrix	QC	Split Ratio	2

Approved:  
T.Patterson  
--e-signature--  
19-Aug-2019

<b>Run Information</b>	<b>Run 1</b>
Filename	7-190816A08
Run Date	16-Aug-19 03:04
Final Volume	10 uL
Dilution Factor	1
Analysis Units	%
Instrument - Column	HRMS-7 DB5MSUSR826231H

Target Analytes	pg	Ret. Time	% Rec	Limits	Flags
2,3,7,8-TCDD	400	28.18	105	70-130	
1,2,3,7,8-PeCDD	2000	32.18	106	70-130	
1,2,3,4,7,8-HxCDD	2000	34.23	102	70-130	
1,2,3,6,7,8-HxCDD	2000	34.28	109	70-130	
1,2,3,7,8,9-HxCDD	2000	34.41	117	70-130	
1,2,3,4,6,7,8-HpCDD	2000	35.90	102	70-130	
OCDD	4000	37.40	97	70-130	
2,3,7,8-TCDF	400	27.26	102	70-130	
1,2,3,7,8-PeCDF	2000	31.26	109	70-130	
2,3,4,7,8-PeCDF	2000	31.97	101	70-130	
1,2,3,4,7,8-HxCDF	2000	33.74	102	70-130	
1,2,3,6,7,8-HxCDF	2000	33.81	115	70-130	
2,3,4,6,7,8-HxCDF	2000	34.13	111	70-130	
1,2,3,7,8,9-HxCDF	2000	34.55	112	70-130	
1,2,3,4,6,7,8-HpCDF	2000	35.34	103	70-130	
1,2,3,4,7,8,9-HpCDF	2000	36.15	101	70-130	
OCDF	4000	37.49	98	70-130	
<b>Field Spike Standards</b>					
37C14-2,3,7,8-TCDD			NS		
13C12-1,2,3,4,7,8-HxCDD			NS		
13C12-2,3,4,7,8-PeCDF			NS		
13C12-1,2,3,4,7,8-HxCDF			NS		
13C12-1,2,3,4,7,8,9-HpCDF			NS		
<b>Extraction Standards</b>					
13C12-2,3,7,8-TCDD	4000	28.17	69	40-130	
13C12-1,2,3,7,8-PeCDD	4000	32.17	93	40-130	
13C12-1,2,3,6,7,8-HxCDD	4000	34.27	77	40-130	
13C12-1,2,3,4,6,7,8-HpCDD	4000	35.89	100	25-130	
13C12-OCDD	8000	37.39	102	25-130	
13C12-2,3,7,8-TCDF	4000	27.25	80	40-130	
13C12-1,2,3,7,8-PeCDF	4000	31.25	97	40-130	
13C12-1,2,3,6,7,8-HxCDF	4000	33.80	82	40-130	
13C12-1,2,3,4,6,7,8-HpCDF	4000	35.33	104	25-130	
<b>Cleanup Standard</b>					
13C12-1,2,3,7,8,9-HxCDF	4000	34.54	89	40-130	

NS Indicates that this standard was not spiked to sample



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

## SVOC DATA PACKAGE

### SECTION 6: INTERNAL RECORDS

Including:

- Prep Logs
- Independent calculation checks
- Others as listed below:

---

---

---

---

---

---

---

---

---

---

14

# Extraction Workup Sheet

Batch ID: WG3116369

Analysis: PUF5 - M23/1668A (HR)

WG3116369

Prep Procedure: BU-TM-1110 Overall HR Prep, BU-TP-1101 8270D Prep, BU-TP-2100 PAH Prep Method

Analyst: Jackson Peagy

Date: 12-Aug-19

## SUBSAMPLING

Sample I.D.	Client I.D.	Media Prep L#
WG3116369-1	Method Blank	L2294735-7
WG3116369-2	Laboratory Control Sample	L2294735-8
WG3116369-3	Extraction and Injection STD.	---
WG3116369-4	Method Blank(Reagent)	Empty Vial
L2310263-1	HEISER-25764410	L2294735-6
L2310263-2	CITY-45764407	L2294735-3
L2310263-3	RES-25764409	L2294735-3

## BATCH TRACKING

	Date/Time/Initials
Client Labels Checked:	12-Aug-19 2:00 PM JP
Media transfer to Soxhlet:	JP
	---
	---
	---
Soxhlet Start Time:	JAZ 12-AUG-19 2:45 PM
Soxhlets Reflux Properly:	JAZ
Soxhlet End Time:	13-Aug-19 9:30 AM EF
	---
	---
	---
Rotovap Reduction + verify temp:	13-Aug-19 AR

JP 12-Aug-19 Extract split:	13-Aug-19 AR
Acid Silica Column:	13-AUG-2019 C11
Solvent exchange:	14-AUG-2019 1600 BS
Alumina Column:	14-AUG-2019 1630 BS
CONDOR COLUMN	15-AUG-19 1500 BS
DX MICRO VIAL	15-AUG-19 1930 BS
Micro/Robo Vial:	14-AUG-2019 1930 BS
Update to LIMS:	14-AUG-19 2:00 BS

\* see comments  
13-Aug-19 AR

Batch ID: WG3116369

**DX Extraction Standard:**

(Checkmark)

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	40	✓
WG3116369-2	40	✓
WG3116369-3	40	✓
WG3116369-4	40	✓
L2310263-1	40	✓
L2310263-2	40	✓
L2310263-3	40	✓
L2316030-1	40	✓
L2316030-2	40	✓
L2316030-3	40	✓
L2320141-1	40	✓
L2320141-2	40	✓
L2320141-3	40	✓
L2322808-1	40	✓
<del>L2322808-2</del>	<del>40</del>	<del>✓</del>
L2322808-3	40	✓

Syringe ID:

320

Standard:

M23-ES#2-035C<sup>SP</sup> 12-Aug-19  
035C

Spike Date:

12-Aug-2019

**Spike Witnessing**

Chemist's Initials

Chemist: JAZ

Witness's Initials

Witness: SP

Witness's Initials

Correct Syringe Obtained: SP

Witness's Initials

Correct Standard Obtained: SP

Witness's Initials

Correct Technique Followed: SP

**PCB Extraction Standard:**

(Checkmark)

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	40	✓
WG3116369-2	40	✓
WG3116369-3	40	✓
WG3116369-4	40	✓
L2310263-1	40	✓
L2310263-2	40	✓
L2310263-3	40	✓
L2316030-1	40	✓
L2316030-2	40	✓
L2316030-3	40	✓
L2320141-1	40	✓
L2320141-2	40	✓
L2320141-3	40	✓
L2322808-1	40	✓
<del>L2322808-2</del>	<del>40</del>	<del>✓</del>
L2322808-3	40	✓

Syringe ID:

118

Standard:

1668A-ES#2-056F

Spike Date:

12-Aug-19

**Spike Witnessing**

Chemist's Initials

Chemist: JAZ

Witness's Initials

Witness: SP

Witness's Initials

Correct Syringe Obtained: SP

Witness's Initials

Correct Standard Obtained: SP

Witness's Initials

Correct Technique Followed: SP

DXPCB STACK PREP

3-Oct-18 / MSM RS

Page 2 of 11

ALS Canada Ltd.

L2310263 DX DPKG 191025

40 of 53



**Batch ID:**

**WG3116369**

Batch ID: WG3116369

DX Native Standard:

(Checkmark) Spiked

Sample I.D.	Volume (ul)	Spiked
WG3116369-2	40	✓
WG3116369-3	40	✓

PCB Native Standard:

(Checkmark) Spiked

Sample I.D.	Volume (ul)	Spiked
WG3116369-2	40	✓
WG3116369-3	40	✓

DX Cleanup Standard:

(Checkmark) Spiked

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	20	✓
WG3116369-2	20	✓
WG3116369-3	N/A	N/A
WG3116369-4	20	✓
L2310263-1	20	✓
L2310263-2	20	✓
L2310263-3	20	✓
L2316030-1	20	✓
L2316030-2	20	✓
L2316030-3	20	✓
L2320141-1	20	✓
L2320141-2	20	✓
L2320141-3	20	✓
L2322808-1	20	✓
<del>L2322808-2</del>	<del>20</del>	<del>✓</del>
L2322808-3	20	✓

PCB Cleanup Standard:

(Checkmark) Spiked

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	20	✓
WG3116369-2	20	✓
WG3116369-3	N/A	N/A
WG3116369-4	20	✓
L2310263-1	20	✓
L2310263-2	20	✓
L2310263-3	20	✓
L2316030-1	20	✓
L2316030-2	20	✓
L2316030-3	20	✓
L2320141-1	20	✓
L2320141-2	20	✓
L2320141-3	20	✓
L2322808-1	20	✓
<del>L2322808-2</del>	<del>20</del>	<del>✓</del>
L2322808-3	20	✓

Syringe ID: 322

Standard: 1613B-NS#3-023E

Date & Initials: 12-Aug-2019 JAZ

Syringe ID: 323

Standard: 1668A-NS#1-36C

Date & Initials: 12-Aug-2019 JAZ

Syringe ID: 357

Standard: M23-CL#1-033B

Date & Initials: 13-AUG-2019 CM

Correct Syringe Obtained:

Chemist's Initials

CM

Correct Standard Obtained:

Chemist's Initials

CM

Correct Technique Followed:

Chemist's Initials

CM

JP 12-Aug-19

Syringe ID: 378

Standard: 1668A-CL#2-32J

Date & Initials: 13-AUG-2019 CM

Correct Syringe Obtained:

Chemist's Initials

CM

Correct Standard Obtained:

Chemist's Initials

CM

Correct Technique Followed:

Chemist's Initials

CM

JP 12-Aug-19

Batch ID: WG3116369

DX Injection Standard:

(Checkmark)

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	10	✓
WG3116369-2	10	✓
WG3116369-3	10	✓
WG3116369-4	10	✓
L2310263-1	10	✓
L2310263-2	10	✓
L2310263-3	10	✓
L2316030-1	10	✓
L2316030-2	10	✓
L2316030-3	10	✓
L2320141-1	10	✓
L2320141-2	10	✓
L2320141-3	10	✓
L2322808-1	10	✓
<del>L2322808-2</del>	<del>10</del>	<del>✓</del>
L2322808-3	10	✓

Syringe ID:

335

Standard:

1613B-IS#1- 0765

Date & Initials:

15-AUG-2019 JB

Correct Syringe Obtained:

Chemist's Initials  
JB

Correct Standard Obtained:

Chemist's Initials  
JB

Correct Technique Followed:

Chemist's Initials  
JB

JP  
12-Aug-19

PCB Injection Standard:

(Checkmark)

Sample I.D.	Volume (ul)	Spiked
WG3116369-1	5	✓
WG3116369-2	5	✓
WG3116369-3	5	✓
WG3116369-4	5	✓
L2310263-1	5	✓
L2310263-2	5	✓
L2310263-3	5	✓
L2316030-1	5	✓
L2316030-2	5	✓
L2316030-3	5	✓
L2320141-1	5	✓
L2320141-2	5	✓
L2320141-3	5	✓
L2322808-1	5	✓
<del>L2322808-2</del>	<del>5</del>	<del>✓</del>
L2322808-3	5	✓

Syringe ID:

260

Standard:

1668A-IS#2- 0114

Date & Initials:

14-AUG-2019 JB

Correct Syringe Obtained:

Chemist's Initials  
JB

Correct Standard Obtained:

Chemist's Initials  
JB

Correct Technique Followed:

Chemist's Initials  
JB

JP  
12-Aug-19

<b>Batch ID:</b>	WG3116369
------------------	-----------

**Reagent Lot Numbers:**

Reagent	Lot#	Manufacturer
Acetone	104436	
Hexane	104314	
DCM	104596	
Toluene	104369	
Nonane	ORG-WAKONON- 648	
1:1 DCM:HEX	ORG-DH2- 229	
Sodium Sulphate	ORG-SSU- 2116	
Acid Silica	ORG-ASI- 8760	
Neutral Silica	ORG-NSI- 1039	
Alumina	ORG-ALU- 437	
1% Deactivated Silica	ORG-2%DAS- -	
Chromacarb	ORG-CC- 241	

Batch ID: WG3116369

**Procedure:**

**This batchsheet is a guideline only. Please see test procedure for complete set of instructions.**

**Extraction:**

- For MB and LCS you **must** use blank media - if not available see your Team Lead
- Place the PUF in to a pre-cleaned thimble and transfer to the Soxhlet body.
- Spike with Extraction Standard (plus Native for LCS and ENI).
- Soxhlet extract in DCM for 16 hours (check with team lead or supervisor)

**Rotovap:**

- Rotovap and reduce to ~5mL.
- Transfer to a calibrated c-tube (marked at 1ml, 2ml) with 3x2ml hexane
- Mix well then quantitatively spilt the extract **1/2 DX/PCB 1/2 Archive**

Batch ID: WG3116369

**DX/PCB:**

- Perform Acid Silica column
- Solvent Exchange (reduce to **~50ul**, bulk back up to 1ml Hexane, vortex well.
- Perform Alumina Column:
  - Pre-elute the Alumina Column with 7ml Hexane
  - Place F1 c-tube under the column, then load the sample with 3x1ml Hexane rinses
  - F1 (Archive) 1mL Hexane
  - F2 (DX/PCB) 14mL 1:1 DCM:Hexane
- Split Alumina F2 **1/2 PCB 1/2 DX**

**Micro-Vial:**

**PCB:**

- Blow down to ~1/2ml
- Vortex **very** well.
- Transfer every last drop to a micro-vial (Marked at 20uL with nonane).
- Blow down to the line
- Spike PCB Injection Standard. cap and vortex. **FV=25ul**

<b>DX:</b>
- Solvent Exchange to Hexane (Reduce to Just Dry then bulk back up to 1ml Hexane)
- ChromaCarb: - 4cm of well-packed chroma-carb.
- Pre-elute Carbon with 5ml Hexane
- Transfer with 3x1ml Hexane
- F1 = <b>10ml</b> 1:1 DCM:Hexane (Archive)
- After dripping has stopped Invert Column.
- F2 = 14ml Toluene (DX and PCB)
- After the column has stopped dripping reduce the <b>F2</b> portion down to ~1/2ml.
- Vortex well, then transfer to a micro-vial without rinses.
- Blow the micro-vial down to just-dry.
- Spike with Injection Standard, Cap the micro-vial, and Vortex. <b>FV=10ul</b>

Batch ID: WG3116369

Comments:

**NOTE: Label and Save All Columns including Acid Silica Columns**

LZ322808-2 sample has not arrived yet JP 12-Aug-19  
\* flask contained ~5ml water after extraction, water was removed with  
pipette. 13-Aug-19 BR

Approval of Deviation from Standard Method

Procedure does deviate from Standard Method. (Batch Writer): \_\_\_\_\_  
Approved (Supervisor/Manager): \_\_\_\_\_

. made new W3 revid not locate original. W3 b  
spike only w DX IS. KB 15-Aug-19



WG3116369		Prep Analyst:			
PUFS - M23/1668A (HR)		Date:			
	Very Good	Meets Method Req	Some Outliers	Very Poor	Comments / Was spl/batch sent for rework? Why?
<b>MB</b>					
<b>LCS</b>					
<b>DUP</b>					
<b>ES rec</b>					

# ALS Life Sciences

## Sample Calculation Report

**CS3 RRF Check**

Approved:	<i>T.Patterson</i> --e-signature-- 19-Aug-2019
-----------	--

$$\text{RRF} = \frac{\text{Response of 2,3,7,8-TCDD}}{\text{Response of 13C12-2,3,7,8-TCDD}} \times \frac{\text{Concentration of 13C12-2,3,7,8-TCDD}}{\text{Concentration of TCDD}}$$

Calculated Value	Value from TargetLynx
------------------	-----------------------

$$\text{RRF} = \frac{143318.60}{1136712.10} \times \frac{100}{10} = 1.26 \quad \text{Value from TargetLynx: } 1.26$$

**Calculation of OCDD amount in L2310263-1**

$$\text{pg} = \frac{\text{Response of OCDD}}{\text{Response of 13C12-OCDD}} \times \frac{\text{pg of 13C12-OCDD spiked}}{\text{Mean RRF} * \text{Sample Size}}$$

$$\text{pg} = \frac{466.5}{522988.9} \times \frac{8000}{0.95 * 1.00} = 7.47 \quad \text{Value from TargetLynx: } 7.47$$

**Calculation of 13C12-2,3,7,8-TCDD Recovery in L2310263-1**

$$\% \text{ Recovery} = \frac{\text{Response of 13C12-2,3,7,8-TCDD}}{\text{Response of 13C12-1,2,3,4-TCDD}} \times \frac{\text{pg of 13C12-1,2,3,4-TCDD spiked} * 100}{\text{Mean RRF} * \text{Amount Spiked}}$$

$$\% \text{ Recovery} = \frac{450065.9}{706498.9} \times \frac{2000 * 100}{0.91 * 2000} = 70 \quad \text{Value from TargetLynx: } 70 \%$$



1435 Norjohn Court, Unit 1, Burlington, ON, Canada L7L 0E6

# SVOC DATA PACKAGE

## SECTION 7: SHIPPING/RECEIVING DOCUMENTS

Including:

- Airbills
- Chain-of-Custody Records
- Sample Log-in Sheet(s) - where applicable
- Others as listed below:

---

---

---

---

---

---

---

---

---

---

[ For lab use only ]



# ANALYTICAL REQUEST FORM

L2310263

1.  REGULAR Status

RUSH Status Requested - ADDITIONAL CHARGE  
 RESULTS REQUIRED BY \_\_\_\_\_ DATE \_\_\_\_\_  
 CONTACT ALS PRIOR TO SENDING SAMPLES

2. Date 7/11/19 Purchase Order No. \_\_\_\_\_

3. Company Name : Floyd Snider ALS Project Manager: Ron McLeod

Address: 601 Union St Suite 600

Seattle WA 98101

Person to Contact: Emily Jones

Telephone (719) 292-2078

Fax Telephone ( ) \_\_\_\_\_

E-mail Address: emily.jones@floydsnider.com

Billing Address (if different from above) \_\_\_\_\_

4. Quote No. \_\_\_\_\_ Email quote \_\_\_\_\_

5. Sample Collection

Sampling Site Background Industrial Areas Phase 1

Industrial Process: Background industry

Date of Collection 7/5/19 - 7/11/19

Time Collected 0900

Date of Shipment \_\_\_\_\_

Chain of Custody No.: \_\_\_\_\_

6. How did you first learn about ALS?  
Referred to by SKC / T&B Systems

### 7. REQUEST FOR ANALYSES

Client Sample Number	Matrix*	Sample/Area Volume	ANALYSES REQUESTED - Use method number if known	Units**	Lab Comments
<u>Heiser - 25764410</u>	<u>Lo-Vol PUF tube</u>		<u>PCBs by Method 1668; Dioxins by Method 8290A</u>	<u>1</u>	<u>1</u>
<u>City - 25764407</u>	<u>Lo-Vol PUF tube</u>		<u>PCBs by Method 1668; Dioxins by Method 8290A</u>	<u>1</u>	<u>2</u>
<u>Res - 25764409</u>	<u>Lo-Vol PUF tube</u>		<u>PCBs by Method 1668; Dioxins by Method 8290A</u>	<u>1</u>	<u>3</u>

\* Specify: Solid sorbent tube, e.g. Charcoal; Filter type; Impinger solution; Bulk sample; Blood; Urine; Tissue; Soil; Water; Other

\*\* 1. µg/sample 2. mg/m<sup>3</sup> 3. ppm 4. % 5. µg/m<sup>3</sup> 6. \_\_\_\_ (other) Please indicate one or more units in the column entitled Units\*\*

Comments HOLD FOR BATCH ANALYSIS AFTER REC'E 5 SAMPLE GROUPS

Possible Contamination and/or Chemical Hazards \_\_\_\_\_

### 7. Chain of Custody (Optional)

Relinquished by <u>Emily Jones</u>	<u>EMILY JONES</u>	Date/Time <u>7/15/19 0900</u>
Received by <u>AARON BURTAU</u>	<u>AARON BURTAU</u>	Date/Time <u>16-JULY-2019 10:40 6.8°C</u>
Relinquished by _____	_____	Date/Time _____
Received by _____	_____	Date/Time _____

1435 Norjohn Court #1 / Burlington, ON Canada, L7L 0E6 800-356-9135 or 801-266-7700 / FAX: 801-268-9992  
 ALS Environmental

## Sample Receiving Log

Date/Time Received	Client ID	Number/Description of Containers	Temp. on Receipt*	Condition of Samples, Courier & Tracking Information	Receiver's Initials	Date/Time Login Completed	Submission ID	Sample ID Range
18-July-2019 10:40	Floyd Snider	3 x air samples	6.8°C	Good FedEx 7884 7124 2481	NZ	18-July-2019 11:15	L2310263	-1-3

\*Temperatures were recorded using:  'Oakton infraPro' dedicated I.R. gun (serial #97800270)  
 Other (specify): .....